



Analytical Summary Report

Analysis Method: 9041A

Analyst By : Iwona

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB124108

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.02	05/15/2023	16:00
2	CAL2	1	Water	NA	NA	20.8	7.00	02/15/2023	16:01
3	CAL3	1	Water	NA	NA	20.9	10.02	02/15/2023	16:02
4	ICV	1	Water	NA	NA	20.9	7.01	02/15/2023	16:04
5	CCV1	1	Water	NA	NA	20.8	2.01	02/15/2023	16:30
6	O1541-01	1	Solid	20	20	20.9	7.01	02/15/2023	16:32
7	O1541-01DUP	1	Solid	20	20	20.9	7.02	02/15/2023	16:34
8	CCV2	1	Water	NA	NA	20.8	12.01	02/15/2023	16:35

WORKLIST(Hardcopy Internal Chain)

LB124108

WorkList Name : pH-021523-2

WorkList ID : 167441

Department : Wet-Chemistry

Date : 02-15-2023 11:56:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/13/2023	Solid	O1541-01	pH	Cool 4 deg C	FREE02	F51	ASH	02/13/2023	9041A

Date/Time 02/15/23 15:45
 Raw Sample Received by: 12/01/ CP Sm
 Raw Sample Relinquished by: CP Sm

Date/Time 02/15/23 16:50
 Raw Sample Received by: CP Sm
 Raw Sample Relinquished by: 12/01/