



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124115

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	02/16/2023	08:40
2	O1549-10		1	12	100.00	>212.0	>212.0	02/16/2023	09:10
3	O1549-10DUP		1	12	100.00	>212.0	>212.0	02/16/2023	09:40
4	O1549-11		1	14	100.00	>212.0	>212.0	02/16/2023	10:10
5	O1553-05		1	13	100.00	>212.0	>212.0	02/16/2023	10:40
6	O1561-03		1	14	100.00	>212.0	>212.0	02/16/2023	11:10
7	O1561-04		1	12	100.00	>212.0	>212.0	02/16/2023	11:40
8	O1563-01		1	13	100.00	>212.0	>212.0	02/16/2023	12:10
9	O1572-07		1	14	100.00	>212.0	>212.0	02/16/2023	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

26124115

WorkList Name : fp-2-16

WorkList ID : 167422

Department : Wet-Chemistry

Date : 02-16-2023 08:22:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/21/2023	Water	O1549-10	Flash Point	Cool 4 deg C	PSEG03	G22	1737	02/14/2023	1010B
02/21/2023	Water	O1549-11	Flash Point	Cool 4 deg C	PSEG03	G22	1929-1930	02/14/2023	1010B
02/21/2023	Water	O1553-05	Flash Point	Cool 4 deg C	LIRO01	G31	TWP-06	02/14/2023	1010B
02/22/2023	Water	O1561-03	Flash Point	Cool 4 deg C	PSEG03	G32	BUR-1244	02/15/2023	1010B
02/22/2023	Water	O1561-04	Flash Point	Cool 4 deg C	PSEG03	G32	BUR-1245	02/15/2023	1010B
02/22/2023	Water	O1563-01	Flash Point	Cool 4 deg C	PSEG03	G32	AUD-1928	02/15/2023	1010B
02/20/2023	Water	O1572-07	Flash Point	Cool 4 deg C	PSEG03	G32	TOTE	02/15/2023	1010B

Date/Time 02/16/2023 08:30
 Raw Sample Received by: RM CP Sm
 Raw Sample Relinquished by: CP Sm

Date/Time 02/16/2023 13:20
 Raw Sample Received by: CP Sm
 Raw Sample Relinquished by: RM CWC