

LB12412,


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

42712



A= 0.0000e+000

B= 2.3407e-004

C= -1.0370e-001

Rho= 0.9995955

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/0D
0.0	0.000	-0.014	-0.014	384	0.000	384					1
0.2	0.200	0.207	0.007	1329	0.0 %	1329					3
2.5	2.500	2.528	0.028	11245	0.0 %	11245					1
5.0	5.000	4.883	-0.117	21306	0.0 %	21306					2
7.5	7.500	7.701	0.201	33341	0.0 %	33341					3
10.0	10.000	9.894	-0.106	42712	0.0 %	42712					1

LB124121 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Feb 2023 08:34:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	384	-	0.0000	7471B	PPB	17 Feb 2023 09:10:33	S	Std
0.2 - 1	50.2	1329	-	0.2000	7471B	PPB	17 Feb 2023 09:12:56	S	Std
2.5 - 1	52.5	11245	-	2.5000	7471B	PPB	17 Feb 2023 09:15:13	S	Std
5.0 - 1	55	21306	-	5.0000	7471B	PPB	17 Feb 2023 09:17:30	S	Std
7.5 - 1	57.5	33341	-	7.5000	7471B	PPB	17 Feb 2023 09:19:47	S	Std
10.0 - 1	511	42712	-	10.0000	7471B	PPB	17 Feb 2023 09:24:36	S	Std
ICV50 - 1	ICV50	18704	4.2744	-	7471B	PPB	17 Feb 2023 09:27:22	U	SMPL
ICB50 - 1	ICB50	404	-0.0091	-	7471B	PPB	17 Feb 2023 09:29:37	U	SMPL
CCV44 - 1	CCV44	21455	4.9183	-	7471B	PPB	17 Feb 2023 09:31:59	U	SMPL
CCB44 - 1	CCB44	440	-0.0007	-	7471B	PPB	17 Feb 2023 09:34:14	U	SMPL
CRA - 1	CRA	1317	0.2046	-	7471B	PPB	17 Feb 2023 09:36:29	U	SMPL
HighStd - 1	HighStd	42444	9.8313	-	7471B	PPB	17 Feb 2023 09:38:45	U	SMPL
ChkStd - 1	ChkStd	30214	6.9686	-	7471B	PPB	17 Feb 2023 09:41:04	U	SMPL
PB150897BL - 1	PBS	187	-0.0599	-	7471B	PPB	17 Feb 2023 09:43:20	U	SMPL
PB150897BS - 1	LCSS	16863	3.8435	-	7471B	PPB	17 Feb 2023 09:45:37	U	SMPL
O1552-01 - 1	SB-01-0-2.0	12172	2.7454	-	7471B	PPB	17 Feb 2023 09:47:54	U	SMPL
O1552-01DUP - 1	SB-01-0-2.0DUP	13264	3.0010	-	7471B	PPB	17 Feb 2023 09:50:11	U	SMPL
O1552-01MS - 1	SB-01-0-2.0MS	30999	7.1523	-	7471B	PPB	17 Feb 2023 09:52:29	U	SMPL
O1552-01MSD - 1	SB-01-0-2.0MSD	29887	6.8920	-	7471B	PPB	17 Feb 2023 09:54:46	U	SMPL
O1552-02 - 1	SB-01-4.0-6.0	5451	1.1722	-	7471B	PPB	17 Feb 2023 09:57:02	U	SMPL
CCV45 - 1	CCV45	20917	4.7924	-	7471B	PPB	17 Feb 2023 09:59:17	U	SMPL
CCB45 - 1	CCB45	-65	-0.1189	-	7471B	PPB	17 Feb 2023 10:01:32	U	SMPL
O1552-05 - 1	SB-03-0-2.0	20311	4.6506	-	7471B	PPB	17 Feb 2023 10:06:03	U	SMPL
O1552-06 - 1	SB-03-4.0-6.0	8299	1.8389	-	7471B	PPB	17 Feb 2023 10:08:18	U	SMPL
O1552-08 - 1	SB-05-0-2.0	8142	1.8021	-	7471B	PPB	17 Feb 2023 10:10:34	U	SMPL
O1552-09 - 1	SB-05-4.0-6.0	3456	0.7053	-	7471B	PPB	17 Feb 2023 10:12:50	U	SMPL
O1552-11 - 1	SB-06-0-2.0	9140	2.0357	-	7471B	PPB	17 Feb 2023 10:15:07	U	SMPL
O1552-12 - 1	SB-06-4.0-6.0	6810	1.4903	-	7471B	PPB	17 Feb 2023 10:17:23	U	SMPL
O1552-14 - 1	SB-04-0-2.0	2598	0.5044	-	7471B	PPB	17 Feb 2023 10:19:41	U	SMPL
O1552-15 - 1	SB-04-4.0-6.0	3209	0.6474	-	7471B	PPB	17 Feb 2023 10:21:58	U	SMPL
O1552-17 - 1	SB-02-0-2.0	8035	1.7771	-	7471B	PPB	17 Feb 2023 10:24:13	U	SMPL
CCV46 - 1	CCV46	20877	4.7830	-	7471B	PPB	17 Feb 2023 10:26:29	U	SMPL
CCB46 - 1	CCB46	15	-0.1002	-	7471B	PPB	17 Feb 2023 10:28:45	U	SMPL
O1552-18 - 1	SB-02-4.0-6.0	7116	1.5620	-	7471B	PPB	17 Feb 2023 10:31:01	U	SMPL
O1555-01 - 1	OR-3-2142023	3311	0.6713	-	7471B	PPB	17 Feb 2023 10:33:16	U	SMPL
O1562-01 - 1	BU-3-021523	2445	0.4686	-	7471B	PPB	17 Feb 2023 10:35:32	U	SMPL
O1570-01 - 1	AU-6-021523	1892	0.3392	-	7471B	PPB	17 Feb 2023 10:37:48	U	SMPL
O1572-01 - 1	COMP-277-281	2284	0.4309	-	7471B	PPB	17 Feb 2023 10:40:06	U	SMPL
O1572-04 - 1	RBR-200028	3126	0.6280	-	7471B	PPB	17 Feb 2023 10:42:22	U	SMPL
O1574-01 - 1	WASTE	2013	0.3675	-	7471B	PPB	17 Feb 2023 10:44:39	U	SMPL
O1552-01LX5 - 1		2508	0.4834	-	7471B	PPB	17 Feb 2023 10:46:57	U	SMPL
O1552-01A - 1		29436	6.7865	-	7471B	PPB	17 Feb 2023 10:49:14	U	SMPL
O1552-03 - 1	SB-01-4.0-6.0-DUP	6063	1.3155	-	7471B	PPB	17 Feb 2023 10:51:29	U	SMPL
CCV47 - 1	CCV47	22143	5.0794	-	7471B	PPB	17 Feb 2023 10:58:59	U	SMPL
CCB47 - 1	CCB47	254	-0.0442	-	7471B	PPB	17 Feb 2023 11:01:14	U	SMPL