



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124125

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	02/17/2023	08:40
2	O1575-02		1	12	100.00	>212.0	>212.0	02/17/2023	09:10
3	O1575-02DUP		1	12	100.00	>212.0	>212.0	02/17/2023	09:40
4	O1579-02		1	14	100.00	>212.0	>212.0	02/17/2023	10:10
5	O1580-01		1	12	100.00	>212.0	>212.0	02/17/2023	10:40
6	O1591-01		1	14	100.00	>212.0	>212.0	02/17/2023	11:10
7	O1591-02		1	12	100.00	>212.0	>212.0	02/17/2023	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6124125

WorkList Name : fp-2-17

WorkList ID : 167456

Department : Wet-Chemistry

Date : 02-17-2023 08:18:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/21/2023	Water	O1575-02	Flash Point	Cool 4 deg C	PSEG03	G41	NWB-1890	02/16/2023	1010B
02/23/2023	Water	O1579-02	Flash Point	Cool 4 deg C	PSEG03	G41	188	02/16/2023	1010B
02/23/2023	Water	O1580-01	Flash Point	Cool 4 deg C	PSEG03	G41	BB-WATERCOMP-0216	02/16/2023	1010B
02/23/2023	Water	O1591-01	Flash Point	Cool 4 deg C	PSEG03	G42	T2908	02/16/2023	1010B
02/23/2023	Water	O1591-02	Flash Point	Cool 4 deg C	PSEG03	G42	T2911-2914	02/16/2023	1010B

Date/Time 02/17/2023 08:20
 Raw Sample Received by: RSM Cures
 Raw Sample Relinquished by: CP Sim

Date/Time 02/17/2023 12:20
 Raw Sample Received by: CP Sim
 Raw Sample Relinquished by: RSM Cures