

LB124130


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

30852



A= 0.0000e+000

B= 3.1816e-004

C= -5.2593e-002

Rho= 0.9988141

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.041	0.041	293	0.000	293					-
0.2	0.200	0.197	-0.003	784	0.0 %	784					-2
2.5	2.500	2.287	-0.213	7353	0.0 %	7353					-4
5.0	5.000	5.154	0.154	16366	0.0 %	16366					3
7.5	7.500	7.758	0.258	24550	0.0 %	24550					3
10.0	10.000	9.763	-0.237	30852	0.0 %	30852					-2

LB124130 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Feb 2023 13:46:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	293	-	0.0000	7470A	PPB	17 Feb 2023 14:00:21	S	Std
0.2 - 1	50.2	784	-	0.2000	7470A	PPB	17 Feb 2023 14:02:39	S	Std
2.5 - 1	52.5	7353	-	2.5000	7470A	PPB	17 Feb 2023 14:04:56	S	Std
5.0 - 1	55	16366	-	5.0000	7470A	PPB	17 Feb 2023 14:07:13	S	Std
7.5 - 1	57.5	24550	-	7.5000	7470A	PPB	17 Feb 2023 14:25:21	S	Std
10.0 - 1	51	30852	-	10.0000	7470A	PPB	17 Feb 2023 14:27:36	S	Std
ICV51 - 1	ICV51	13391	4.2078	-	7470A	PPB	17 Feb 2023 14:38:24	U	SMPL
ICB51 - 1	ICB51	192	0.0085	-	7470A	PPB	17 Feb 2023 14:40:40	U	SMPL
CCV48 - 1	CCV48	16433	5.1757	-	7470A	PPB	17 Feb 2023 14:48:17	U	SMPL
CCB48 - 1	CCB48	379	0.0680	-	7470A	PPB	17 Feb 2023 14:50:32	U	SMPL
CRA - 1	CRA	727	0.1787	-	7470A	PPB	17 Feb 2023 14:52:47	U	SMPL
HighStd - 1	HighStd	30709	9.7177	-	7470A	PPB	17 Feb 2023 14:55:03	U	SMPL
ChkStd - 1	ChkStd	22853	7.2183	-	7470A	PPB	17 Feb 2023 14:57:18	U	SMPL
PB150908BL - 1	PBW	175	0.0031	-	7470A	PPB	17 Feb 2023 14:59:34	U	SMPL
PB150908BS - 1	LCSW	11386	3.5699	-	7470A	PPB	17 Feb 2023 15:01:50	U	SMPL
O1552-04 - 1	SB-01-COMP	213	0.0152	-	7470A	PPB	17 Feb 2023 15:04:06	U	SMPL
O1552-07 - 1	SB-03-COMP	302	0.0435	-	7470A	PPB	17 Feb 2023 15:06:23	U	SMPL
O1552-10 - 1	SB-05-COMP	358	0.0613	-	7470A	PPB	17 Feb 2023 15:08:40	U	SMPL
O1552-13 - 1	SB-06-COMP	396	0.0734	-	7470A	PPB	17 Feb 2023 15:10:57	U	SMPL
O1552-16 - 1	SB-04-COMP	215	0.0158	-	7470A	PPB	17 Feb 2023 15:13:12	U	SMPL
CCV49 - 1	CCV49	17251	5.4359	-	7470A	PPB	17 Feb 2023 15:15:30	U	SMPL
CCB49 - 1	CCB49	220	0.0174	-	7470A	PPB	17 Feb 2023 15:17:45	U	SMPL
O1552-19 - 1	SB-02-COMP	112	-0.0170	-	7470A	PPB	17 Feb 2023 15:20:00	U	SMPL
O1572-02 - 1	COMP-277-281	212	0.0149	-	7470A	PPB	17 Feb 2023 15:22:15	U	SMPL
O1572-02DUP - 1	COMP-277-281DUP	233	0.0215	-	7470A	PPB	17 Feb 2023 15:24:31	U	SMPL
O1572-02MS - 1	COMP-277-281MS	11401	3.5747	-	7470A	PPB	17 Feb 2023 15:26:47	U	SMPL
O1572-05 - 1	RBR-200028	203	0.0120	-	7470A	PPB	17 Feb 2023 15:31:20	U	SMPL
O1574-02 - 1	WASTE	261	0.0304	-	7470A	PPB	17 Feb 2023 15:33:36	U	SMPL
PB150882TB - 1	PB150882TB	313	0.0470	-	7470A	PPB	17 Feb 2023 15:35:53	U	SMPL
O1572-02LX5 - 1		107	-0.0186	-	7470A	PPB	17 Feb 2023 15:47:28	U	SMPL
O1572-02A - 1		12335	3.8719	-	7470A	PPB	17 Feb 2023 15:49:44	U	SMPL
CCV50 - 1	CCV50	16579	5.2221	-	7470A	PPB	17 Feb 2023 15:52:00	U	SMPL
CCB50 - 1	CCB50	239	0.0234	-	7470A	PPB	17 Feb 2023 15:54:16	U	SMPL
O1572-02MSD - 1	COMP-277-281MSD	12853	4.0367	-	7470A	PPB	17 Feb 2023 15:59:07	U	SMPL
CCV51 - 1	CCV51	15708	4.9450	-	7470A	PPB	17 Feb 2023 16:04:00	U	SMPL
CCB51 - 1	CCB51	414	0.0791	-	7470A	PPB	17 Feb 2023 16:06:16	U	SMPL