



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Iwona

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB124135

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.00	02/17/2023	12:50
2	CAL2	1	Water	NA	NA	21.0	7.02	02/17/2023	12:51
3	CAL3	1	Water	NA	NA	21.0	10.02	02/17/2023	12:52
4	ICV	1	Water	NA	NA	20.9	7.01	02/17/2023	12:54
5	CCV1	1	Water	NA	NA	20.09	2.01	02/17/2023	14:00
6	O1562-01	1	Solid	20	20	21.2	6.48	02/17/2023	14:05
7	O1562-01DUP	1	Solid	20	20	21.2	6.49	02/17/2023	14:07
8	O1570-01	1	Solid	20	20	21.0	8.47	02/17/2023	14:10
9	CCV2	1	Water	NA	NA	21.1	12.03	02/17/2023	14:15

WORKLIST(Hardcopy Internal Chain)

L B 124135

WorkList Name : pH-021723 WorkList ID : 167474 Department : Wet-Chemistry Date : 02-17-2023 10:07:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/15/2023	Solid	O1562-01	pH	Cool 4 deg C	PSEG05	G32	BU-3-021523	02/15/2023	9045D
02/15/2023	Solid	O1570-01	pH	Cool 4 deg C	PSEG05	G41	AU-6-021523	02/15/2023	9045D

Date/Time 02/17/23 12:30
Raw Sample Received by: 12 (SC)
Raw Sample Relinquished by: CP SM

Date/Time 02/17/23 14:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: 12 (SC)