



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/21/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 02/20/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 02/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB124145

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1601-01	EXHD0	1	1.12	8.42	9.54	8.58	88.6	
O1601-02	EXHD1	2	1.13	8.56	9.69	8.48	85.9	
O1601-03	EXHD2	3	1.12	8.72	9.84	8.56	85.3	
O1601-04	EXHD3	4	1.13	8.43	9.56	8.37	85.9	
O1601-05	EXHD4	5	1.14	8.66	9.8	8.96	90.3	
O1601-06	EXHD5	6	1.15	8.54	9.69	7.67	76.3	
O1601-07	EXHD6	7	1.15	8.51	9.66	7.97	80.1	
O1601-08	EXHD7	8	1.14	8.55	9.69	8.18	82.3	
O1601-09	EXHD8	9	1.14	8.42	9.56	8.23	84.2	
O1601-10	EXHD8MS	10	1.14	8.42	9.56	8.23	84.2	
O1601-11	EXHD8MSD	11	1.14	8.42	9.56	8.23	84.2	
O1601-12	EXHD9	12	1.16	8.69	9.85	8.21	81.1	
O1601-13	EXHF0	13	1.15	8.82	9.97	8.59	84.4	
O1601-14	EXHF1	14	1.13	8.43	9.56	7.95	80.9	
O1601-15	EXHF2	15	1.16	8.60	9.76	8.5	85.3	
O1601-16	EXHF3	16	1.18	8.47	9.65	8.28	83.8	
O1601-17	EXHF4	17	1.11	8.56	9.67	8.3	84.0	
O1601-18	EXHF5	18	1.12	8.60	9.72	8.44	85.1	
O1601-19	EXHF6	19	1.15	8.69	9.84	7.82	76.8	
O1601-20	EXHF7	20	1.18	8.67	9.85	8.47	84.1	
O1601-21	EXHF8	21	1.18	8.57	9.75	8.42	84.5	
O1601-22	EXHF9	22	1.13	8.44	9.57	8.22	84.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$