



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/21/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:35
In Date: 02/20/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 02/21/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB124154

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1620-01	MH0E56	1	1.14	8.56	9.7	9.48	97.4	
O1620-02	MH0E57	2	1.14	8.52	9.66	9.44	97.4	
O1620-03	MH0E60	3	1.12	8.43	9.55	9.36	97.7	
O1620-04	MH0E61	4	1.13	8.58	9.71	9.52	97.8	
O1620-05	MH0E62	5	1.13	8.51	9.64	9.39	97.1	
O1620-06	MH0E63	6	1.12	8.66	9.78	9.51	96.9	
O1620-07	MH0E66	7	1.13	8.61	9.74	9.51	97.3	
O1620-08	MH0E66D	8	1.13	8.61	9.74	9.51	97.3	
O1620-09	MH0E66S	9	1.13	8.61	9.74	9.51	97.3	
O1620-10	MH0E67	10	1.15	8.78	9.93	9.73	97.7	
O1620-11	MH0E68	11	1.13	8.39	9.52	9.22	96.4	
O1620-12	MH0E71	12	1.13	8.44	9.57	9.29	96.7	
O1620-13	MH0E72	13	1.15	8.47	9.62	9.34	96.7	
O1620-14	MH0E73	14	1.12	8.57	9.69	9.36	96.1	
O1620-15	MH0E76	15	1.14	8.70	9.84	9.42	95.2	
O1620-16	MH0E77	16	1.13	8.72	9.85	9.4	94.8	
O1620-17	MH0E78	17	1.13	8.42	9.55	9.08	94.4	
O1620-18	MH0ED4	18	1.12	8.61	9.73	9.55	97.9	
O1620-19	MH0ED7	19	1.12	8.74	9.86	9.58	96.8	
O1620-20	MH0ED8	20	1.14	8.66	9.8	9.52	96.8	
O1620-21	MH0ED9	21	1.13	8.52	9.65	9.48	98.0	
O1620-22	MH0EE2	22	1.14	8.38	9.52	9.31	97.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$