

LB124161


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33816



A= 0.0000e+000

B= 2.9417e-004

C= 4.9800e-002

Rho= 0.9999204

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.049	0.049	-3	0.000	-3					
0.2	0.200	0.218	0.018	573	0.0 %	573					
2.5	2.500	2.413	-0.087	8032	0.0 %	8032					
5.0	5.000	4.977	-0.023	16750	0.0 %	16750					
7.5	7.500	7.546	0.046	25482	0.0 %	25482					
10.0	10.000	9.997	-0.003	33816	0.0 %	33816					

4.0
9
-3
0
1
0

LB124161 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Feb 2023 16:05:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.0000	7470A	PPB	20 Feb 2023 16:11:17	S	Std
0.2 - 1	51.2	573	-	0.2000	7470A	PPB	20 Feb 2023 16:13:34	S	Std
2.5 - 1	52.5	8032	-	2.5000	7470A	PPB	20 Feb 2023 16:15:52	S	Std
5.0 - 1	55	16750	-	5.0000	7470A	PPB	20 Feb 2023 16:18:08	S	Std
7.5 - 1	57.5	25482	-	7.5000	7470A	PPB	20 Feb 2023 16:20:25	S	Std
10.0 - 1	511	33816	-	10.0000	7470A	PPB	20 Feb 2023 16:27:54	S	Std
ICV54 - 1	ICV54	14455	4.3020	-	7470A	PPB	20 Feb 2023 16:30:35	U	SMPL
ICB54 - 1	ICB54	7	0.0519	-	7470A	PPB	20 Feb 2023 16:32:51	U	SMPL
CCV58 - 1	CCV58	16713	4.9662	-	7470A	PPB	20 Feb 2023 16:35:06	U	SMPL
CCB58 - 1	CCB58	-31	0.0407	-	7470A	PPB	20 Feb 2023 16:37:21	U	SMPL
CRA - 1	CRA	572	0.2181	-	7470A	PPB	20 Feb 2023 16:39:37	U	SMPL
HighStd - 1	HighStd	33193	9.8141	-	7470A	PPB	20 Feb 2023 16:41:52	U	SMPL
ChkStd - 1	ChkStd	23865	7.0701	-	7470A	PPB	20 Feb 2023 16:44:07	U	SMPL
PB150972BL - 1	PBW	-61	0.0319	-	7470A	PPB	20 Feb 2023 16:46:24	U	SMPL
PB150972BS - 1	LCSW	14895	4.4314	-	7470A	PPB	20 Feb 2023 16:48:41	U	SMPL
O1553-06 - 1	TWP-06	2271	0.7179	-	7470A	PPB	20 Feb 2023 16:50:57	U	SMPL
O1553-06DUP - 1	TWP-06DUP	2061	0.6561	-	7470A	PPB	20 Feb 2023 16:53:13	U	SMPL
O1553-06MS - 1	TWP-06MS	16420	4.8800	-	7470A	PPB	20 Feb 2023 16:55:31	U	SMPL
O1553-06MSD - 1	TWP-06MSD	14851	4.4185	-	7470A	PPB	20 Feb 2023 16:57:48	U	SMPL
O1561-04 - 1	BUR-1245	462	0.1857	-	7470A	PPB	20 Feb 2023 17:00:04	U	SMPL
CCV59 - 1	CCV59	18261	5.4216	-	7470A	PPB	20 Feb 2023 17:04:51	U	SMPL
CCB59 - 1	CCB59	-50	0.0351	-	7470A	PPB	20 Feb 2023 17:07:07	U	SMPL
O1572-07 - 1	TOTE	2026	0.6458	-	7470A	PPB	20 Feb 2023 17:09:22	U	SMPL
O1593-07 - 1	SW-2	85	0.0748	-	7470A	PPB	20 Feb 2023 17:11:38	U	SMPL
O1593-09 - 1	SW-3	614	0.2304	-	7470A	PPB	20 Feb 2023 17:13:54	U	SMPL
O1593-10 - 1	SW-1	483	0.1919	-	7470A	PPB	20 Feb 2023 17:16:10	U	SMPL
O1593-11 - 1	EB-2023-2-16	35	0.0601	-	7470A	PPB	20 Feb 2023 17:18:27	U	SMPL
O1597-01 - 1	NYE-WATER-0217	5718	1.7318	-	7470A	PPB	20 Feb 2023 17:20:45	U	SMPL
O1607-01 - 1	DRUM	-66	0.0304	-	7470A	PPB	20 Feb 2023 17:23:01	U	SMPL
O1611-03 - 1	EB-2023-2-17	-73	0.0283	-	7470A	PPB	20 Feb 2023 17:25:19	U	SMPL
O1553-06LX5 - 1		456	0.1839	-	7470A	PPB	20 Feb 2023 17:27:36	U	SMPL
O1553-06A - 1		15420	4.5858	-	7470A	PPB	20 Feb 2023 17:29:51	U	SMPL
CCV60 - 1	CCV60	17345	5.1521	-	7470A	PPB	20 Feb 2023 17:32:07	U	SMPL
CCB60 - 1	CCB60	-71	0.0289	-	7470A	PPB	20 Feb 2023 17:34:23	U	SMPL