



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124162

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	02/20/2023	14:30
2	O1597-01		1	12	100.00	>212.0	>212.0	02/20/2023	15:00
3	O1597-01DUP		1	12	100.00	>212.0	>212.0	02/20/2023	15:30
4	O1607-01		1	14	100.00	>212.0	>212.0	02/20/2023	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

(6124182

WorkList Name : f p 2-20

WorkList ID : 167488

Department : Wet-Chemistry

Date : 02-20-2023 08:00:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/23/2023	Water	O1597-01	Flash Point	Cool 4 deg C	PSEG03	G51	NYE-WATER-0217	02/17/2023	1010B
02/22/2023	Water	O1607-01	Flash Point	Cool 4 deg C	PSEG03	G51	DRUM	02/17/2023	1010B

Date/Time 02/20/2023 14:00
 Raw Sample Received by: RM GWO
 Raw Sample Relinquished by: CP SM

Date/Time 02/20/2023 16:40
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: RM CWC