


LB124165

7470A

INSTRUMENT 11: CVL

Linear



A= 0.0000e+000

B= 2.8558e-004

C= 5.5823e-002

Rho= 0.9997569

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.057	0.057	3	0.000	3					
0.2	0.200	0.222	0.022	582	0.0 %	582					11
2.5	2.500	2.344	-0.156	8013	0.0 %	8013					-6
5.0	5.000	5.102	0.102	17671	0.0 %	17671					2
7.5	7.500	7.466	-0.034	25946	0.0 %	25946					0
10.0	10.000	10.009	0.009	34853	0.0 %	34853					0

LB124165 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Feb 2023 10:27:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	7470A	PPB	21 Feb 2023 10:28:05	S	Std
0.2 - 1	50.2	582	-	0.2000	7470A	PPB	21 Feb 2023 10:30:23	S	Std
2.5 - 1	52.5	8013	-	2.5000	7470A	PPB	21 Feb 2023 10:32:40	S	Std
5.0 - 1	55	17671	-	5.0000	7470A	PPB	21 Feb 2023 10:34:57	S	Std
7.5 - 1	57.5	25946	-	7.5000	7470A	PPB	21 Feb 2023 10:37:14	S	Std
10.0 - 1	60	34853	-	10.0000	7470A	PPB	21 Feb 2023 10:44:40	S	Std
ICV55 - 1	ICV55	13470	3.9026	-	7470A	PPB	21 Feb 2023 10:49:55	U	SMPL
ICB55 - 1	ICB55	2	0.0564	-	7470A	PPB	21 Feb 2023 10:52:10	U	SMPL
CCV61 - 1	CCV61	16585	4.7922	-	7470A	PPB	21 Feb 2023 10:54:25	U	SMPL
CCB61 - 1	CCB61	-4	0.0547	-	7470A	PPB	21 Feb 2023 10:56:40	U	SMPL
CRA - 1	CRA	575	0.2200	-	7470A	PPB	21 Feb 2023 10:58:56	U	SMPL
HighStd - 1	HighStd	34111	9.7973	-	7470A	PPB	21 Feb 2023 11:01:13	U	SMPL
ChkStd - 1	ChkStd	24219	6.9723	-	7470A	PPB	21 Feb 2023 11:03:28	U	SMPL
PB150940BL - 1	PBW	-69	0.0361	-	7470A	PPB	21 Feb 2023 11:05:44	U	SMPL
PB150940BS - 1	LCSW	14824	4.2893	-	7470A	PPB	21 Feb 2023 11:08:00	U	SMPL
O1576-02 - 1	RE137-BAGFILTER-20220215	-44	0.0433	-	7470A	PPB	21 Feb 2023 11:10:16	U	SMPL
O1581-02 - 1	MH-A	-17	0.0510	-	7470A	PPB	21 Feb 2023 11:12:33	U	SMPL
O1581-05 - 1	MH-C	-27	0.0481	-	7470A	PPB	21 Feb 2023 11:14:50	U	SMPL
O1581-05DUP - 1	MH-CDUP	-25	0.0487	-	7470A	PPB	21 Feb 2023 11:17:08	U	SMPL
O1581-05MS - 1	MH-CMS	13426	3.8900	-	7470A	PPB	21 Feb 2023 11:19:23	U	SMPL
CCV62 - 1	CCV62	16813	4.8573	-	7470A	PPB	21 Feb 2023 11:21:39	U	SMPL
CCB62 - 1	CCB62	-36	0.0455	-	7470A	PPB	21 Feb 2023 11:23:54	U	SMPL
O1581-05MSD - 1	MH-CMSD	12063	3.5008	-	7470A	PPB	21 Feb 2023 11:26:10	U	SMPL
PB150941BL - 1	PBW	-6	0.0541	-	7470A	PPB	21 Feb 2023 11:28:25	U	SMPL
PB150941BS - 1	LCSW	12188	3.5365	-	7470A	PPB	21 Feb 2023 11:30:41	U	SMPL
O1579-02 - 1	188	85	0.0801	-	7470A	PPB	21 Feb 2023 11:32:57	U	SMPL
O1580-01 - 1	BB-WATERCOMP-0216	-41	0.0441	-	7470A	PPB	21 Feb 2023 11:35:13	U	SMPL
O1591-01 - 1	T2908	-5	0.0544	-	7470A	PPB	21 Feb 2023 11:37:29	U	SMPL
O1591-02 - 1	T2911-2914	7	0.0578	-	7470A	PPB	21 Feb 2023 11:39:46	U	SMPL
O1591-02DUP - 1	T2911-2914DUP	45	0.0687	-	7470A	PPB	21 Feb 2023 11:42:03	U	SMPL
O1591-02MS - 1	T2911-2914MS	13403	3.8835	-	7470A	PPB	21 Feb 2023 11:44:20	U	SMPL
O1591-02MSD - 1	T2911-2914MSD	13267	3.8446	-	7470A	PPB	21 Feb 2023 11:46:36	U	SMPL
CCV63 - 1	CCV63	17737	5.1212	-	7470A	PPB	21 Feb 2023 11:48:52	U	SMPL
CCB63 - 1	CCB63	41	0.0675	-	7470A	PPB	21 Feb 2023 11:51:08	U	SMPL
PB150973BL - 1	PBW	59	0.0727	-	7470A	PPB	21 Feb 2023 11:53:23	U	SMPL
PB150973BS - 1	LCSW	12419	3.6025	-	7470A	PPB	21 Feb 2023 11:55:39	U	SMPL
O1575-01 - 1	NWB-1887	25	0.0630	-	7470A	PPB	21 Feb 2023 11:57:55	U	SMPL
O1579-01 - 1	186	33	0.0652	-	7470A	PPB	21 Feb 2023 12:00:11	U	SMPL
O1582-02 - 1	LEONIA-TRENCH-LINE	72	0.0764	-	7470A	PPB	21 Feb 2023 12:02:27	U	SMPL
O1586-02 - 1	COMP-1	27	0.0635	-	7470A	PPB	21 Feb 2023 12:04:44	U	SMPL
O1586-04 - 1	COMP-2	54	0.0712	-	7470A	PPB	21 Feb 2023 12:07:00	U	SMPL
O1586-06 - 1	COMP-3	45	0.0687	-	7470A	PPB	21 Feb 2023 12:09:17	U	SMPL
O1586-08 - 1	COMP-4	19	0.0612	-	7470A	PPB	21 Feb 2023 12:11:34	U	SMPL
O1586-10 - 1	COMP-5	-35	0.0458	-	7470A	PPB	21 Feb 2023 12:13:50	U	SMPL
CCV64 - 1	CCV64	17719	5.1161	-	7470A	PPB	21 Feb 2023 12:16:06	U	SMPL
CCB64 - 1	CCB64	-27	0.0481	-	7470A	PPB	21 Feb 2023 12:18:22	U	SMPL
O1606-02 - 1	VNJ-210	0	0.0558	-	7470A	PPB	21 Feb 2023 12:20:38	U	SMPL
O1606-02DUP - 1	VNJ-210DUP	-23	0.0493	-	7470A	PPB	21 Feb 2023 12:22:54	U	SMPL
O1606-02MS - 1	VNJ-210MS	13062	3.7861	-	7470A	PPB	21 Feb 2023 12:25:10	U	SMPL
O1606-02MSD - 1	VNJ-210MSD	13671	3.9600	-	7470A	PPB	21 Feb 2023 12:27:25	U	SMPL
PB150974BL - 1	PBW	-64	0.0375	-	7470A	PPB	21 Feb 2023 12:29:41	U	SMPL
PB150974BS - 1	LCSW	11908	3.4565	-	7470A	PPB	21 Feb 2023 12:31:58	U	SMPL
O1553-02 - 1	TWP-01	16	0.0604	-	7470A	PPB	21 Feb 2023 12:34:15	U	SMPL
O1553-02DUP - 1	TWP-01DUP	57	0.0721	-	7470A	PPB	21 Feb 2023 12:36:32	U	SMPL
O1553-02MS - 1	TWP-01MS	13847	4.0103	-	7470A	PPB	21 Feb 2023 12:38:49	U	SMPL
O1553-02MSD - 1	TWP-01MSD	13965	4.0440	-	7470A	PPB	21 Feb 2023 12:41:05	U	SMPL
CCV65 - 1	CCV65	17421	5.0309	-	7470A	PPB	21 Feb 2023 12:43:21	U	SMPL
CCB65 - 1	CCB65	-42	0.0438	-	7470A	PPB	21 Feb 2023 12:45:37	U	SMPL
O1553-04 - 1	TWP-01-DUP	59	0.0727	-	7470A	PPB	21 Feb 2023 12:47:53	U	SMPL
O1553-07 - 1	TWP-06	4	0.0570	-	7470A	PPB	21 Feb 2023 12:50:08	U	SMPL
PB150904TB - 1	PB150904TB	-12	0.0524	-	7470A	PPB	21 Feb 2023 12:52:24	U	SMPL
PB150918TB - 1	PB150918TB	-24	0.0490	-	7470A	PPB	21 Feb 2023 12:54:40	U	SMPL
PB150924TB - 1	PB150924TB	-6	0.0541	-	7470A	PPB	21 Feb 2023 12:56:56	U	SMPL
CCV66 - 1	CCV66	17604	5.0832	-	7470A	PPB	21 Feb 2023 12:59:12	U	SMPL
CCB66 - 1	CCB66	-76	0.0341	-	7470A	PPB	21 Feb 2023 13:01:29	U	SMPL
O1581-05LX5 - 1		-34	0.0461	-	7470A	PPB	21 Feb 2023 13:03:46	U	SMPL
O1581-05A - 1		13313	3.8578	-	7470A	PPB	21 Feb 2023 13:06:03	U	SMPL
O1591-02LX5 - 1		-49	0.0418	-	7470A	PPB	21 Feb 2023 13:08:20	U	SMPL
O1591-02A - 1		13323	3.8606	-	7470A	PPB	21 Feb 2023 13:10:36	U	SMPL
O1606-02LX5 - 1		-41	0.0441	-	7470A	PPB	21 Feb 2023 13:12:53	U	SMPL
O1606-02A - 1		13536	3.9215	-	7470A	PPB	21 Feb 2023 13:15:09	U	SMPL
O1553-02LX5 - 1		-30	0.0473	-	7470A	PPB	21 Feb 2023 13:17:25	U	SMPL
O1553-02A - 1		13855	4.0126	-	7470A	PPB	21 Feb 2023 13:19:42	U	SMPL
CCV67 - 1	CCV67	17617	5.0869	-	7470A	PPB	21 Feb 2023 13:21:58	U	SMPL
CCB67 - 1	CCB67	-44	0.0433	-	7470A	PPB	21 Feb 2023 13:24:14	U	SMPL