


LB124189

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0594e-004

C= -4.2400e-002

Rho= 0.9999192

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.037	-0.037	24	0.000	24					
0.2	0.200	0.163	-0.037	999	0.0 %	999					-19
2.5	2.500	2.550	0.050	12589	0.0 %	12589					2
5.0	5.000	5.070	0.070	24827	0.0 %	24827					1
7.5	7.500	7.508	0.008	36662	0.0 %	36662					0
10.0	10.000	9.946	-0.054	48502	0.0 %	48502					-1

LB124189 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Feb 2023 13:55:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	7471B	PPB	22 Feb 2023 13:59:15	S	Std
0.2 - 1	50.2	999	-	0.2000	7471B	PPB	22 Feb 2023 14:01:33	S	Std
2.5 - 1	52.5	12589	-	2.5000	7471B	PPB	22 Feb 2023 14:03:50	S	Std
5.0 - 1	55	24827	-	5.0000	7471B	PPB	22 Feb 2023 14:06:07	S	Std
7.5 - 1	57.5	36662	-	7.5000	7471B	PPB	22 Feb 2023 14:15:05	S	Std
10.0 - 1	511	48502	-	10.0000	7471B	PPB	22 Feb 2023 14:19:56	S	Std
ICV58 - 1	ICV58	19700	4.0146	-	7471B	PPB	22 Feb 2023 14:25:13	U	SMPL
ICB58 - 1	ICB58	-17	-0.0459	-	7471B	PPB	22 Feb 2023 14:27:28	U	SMPL
CCV72 - 1	CCV72	25187	5.1445	-	7471B	PPB	22 Feb 2023 14:29:43	U	SMPL
CCB72 - 1	CCB72	-23	-0.0471	-	7471B	PPB	22 Feb 2023 14:31:59	U	SMPL
CRA - 1	CRA	1025	0.1687	-	7471B	PPB	22 Feb 2023 14:34:14	U	SMPL
HighStd - 1	HighStd	49421	10.1352	-	7471B	PPB	22 Feb 2023 14:36:29	U	SMPL
ChkStd - 1	ChkStd	35187	7.2039	-	7471B	PPB	22 Feb 2023 14:38:44	U	SMPL
PB151010BL - 1	PB151010BL	-74	-0.0576	-	7471B	PPB	22 Feb 2023 14:41:00	U	SMPL
PB151010BS - 1	PB151010BS	19228	3.9174	-	7471B	PPB	22 Feb 2023 14:48:11	U	SMPL
O1636-01 - 1	OR-02-022123	1727	0.3133	-	7471B	PPB	22 Feb 2023 14:50:27	U	SMPL
O1638-01 - 1	VB-15-241	16269	3.3080	-	7471B	PPB	22 Feb 2023 14:52:43	U	SMPL
O1639-01 - 1	TR-05-02212023	1811	0.3306	-	7471B	PPB	22 Feb 2023 14:55:01	U	SMPL
O1640-01 - 1	TP-1	1540	0.2747	-	7471B	PPB	22 Feb 2023 14:57:17	U	SMPL
O1640-01DUP - 1	TP-1DUP	1538	0.2743	-	7471B	PPB	22 Feb 2023 14:59:33	U	SMPL
CCV73 - 1	CCV73	25193	5.1458	-	7471B	PPB	22 Feb 2023 15:01:48	U	SMPL
CCB73 - 1	CCB73	-39	-0.0504	-	7471B	PPB	22 Feb 2023 15:04:04	U	SMPL
O1640-01MS - 1	TP-1MS	19517	3.9769	-	7471B	PPB	22 Feb 2023 15:06:19	U	SMPL
O1640-01MSD - 1	TP-1MSD	20596	4.1991	-	7471B	PPB	22 Feb 2023 15:08:35	U	SMPL
O1640-02 - 1	TP-2	1037	0.1712	-	7471B	PPB	22 Feb 2023 15:10:50	U	SMPL
O1640-03 - 1	TP-3	7258	1.4523	-	7471B	PPB	22 Feb 2023 15:13:06	U	SMPL
O1640-04 - 1	TP-4	779	0.1180	-	7471B	PPB	22 Feb 2023 15:15:22	U	SMPL
O1640-01LX5 - 1		284	0.0161	-	7471B	PPB	22 Feb 2023 15:17:39	U	SMPL
O1640-01A - 1		21392	4.3630	-	7471B	PPB	22 Feb 2023 15:19:56	U	SMPL
O1418-03DLX5 - 1		24463	4.9954	-	7471B	PPB	22 Feb 2023 15:27:10	U	SMPL
O1418-03DUPDLX5 - 1		28039	5.7319	-	7471B	PPB	22 Feb 2023 15:29:26	U	SMPL
O1418-03MSDLX5 - 1		30397	6.2175	-	7471B	PPB	22 Feb 2023 15:31:42	U	SMPL
CCV74 - 1	CCV74	25322	5.1723	-	7471B	PPB	22 Feb 2023 15:33:58	U	SMPL
CCB74 - 1	CCB74	-22	-0.0469	-	7471B	PPB	22 Feb 2023 15:36:14	U	SMPL
O1418-03MSDDLX5 - 1		31364	6.4166	-	7471B	PPB	22 Feb 2023 15:38:29	U	SMPL
O1418-03ADLX5 - 1		31062	6.3544	-	7471B	PPB	22 Feb 2023 15:42:22	U	SMPL
O1418-03LDLX25 - 1		4462	0.8765	-	7471B	PPB	22 Feb 2023 15:49:57	U	SMPL
O1418-11DLX10 - 1		21955	4.4789	-	7471B	PPB	22 Feb 2023 15:54:28	U	SMPL
O1418-07DLX5 - 1		16192	3.2921	-	7471B	PPB	22 Feb 2023 16:04:57	U	SMPL
O1418-15DLX5 - 1		15834	3.2184	-	7471B	PPB	22 Feb 2023 16:09:30	U	SMPL
CCV75 - 1	CCV75	25120	5.1307	-	7471B	PPB	22 Feb 2023 16:11:46	U	SMPL
CCB75 - 1	CCB75	-59	-0.0545	-	7471B	PPB	22 Feb 2023 16:14:03	U	SMPL