

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124193

Reagent/Standard	Lot/Log #
COD calibration std. 10 ppm	WP100109
COD calibration std. 150 ppm	WP100112
COD calibration std. 50 ppm	WP100110
COD calibration std. 0 ppm	WP100108
COD ICV-LCS std, 50ppm	WP100114
COD calibration std. 100 ppm	WP100111
COD Digestion Vials Low Level 0-150Mg/L	W2997
COD ICV-LCS std, 50ppm	WP101124
COD CCV std, 50ppm	WP101123
COD LOD std, 5ppm	WP101125
COD LOQ std, 10.0PPM	WP101126

Temp In (C): 148	Date In: 02/22/2023	Time In: 09:15
Temp Out (C): 150	Date Out: 02/22/2023	Time Out: 11:15

Intercept: -0.039

Slope: 1.0006

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	0.039		12/05/2022	15:25
2	CAL2	10	1	Water	11.000	11.032	10.3	12/05/2022	15:25
3	CAL3	50	1	Water	49.000	49.01	-2	12/05/2022	15:26
4	CAL4	100	1	Water	99.000	98.98	-1	12/05/2022	15:26
5	CAL5	150	1	Water	151.000	150.948	0.6	12/05/2022	15:27



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124193

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	51.000	51.008	12/05/2022	15:27
2	ICB		NA	NA	1	Water	1.000	1.038	12/05/2022	15:28
3	CCV1	50	NA	NA	1	Water	51.000	51.008	02/22/2023	14:45
4	CCB1		NA	NA	1	Water	1.000	1.038	02/22/2023	14:45
5	LB124193BL		1	50	1	Solid	1.000	1.038	02/22/2023	14:46
6	LB124193BS	50	1	50	1	Solid	51.000	51.008	02/22/2023	14:46
7	O1627-01		1	50	1	Solid	6.000	6.035	02/22/2023	14:47
8	O1627-02		1	50	1	Solid	10.000	10.033	02/22/2023	14:47
9	CCV2	50	NA	NA	1	Water	51.000	51.008	02/22/2023	14:48
10	CCB2		NA	NA	1	Water	0.000	0.039	02/22/2023	14:48