



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124202

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	02/22/2023	16:20
2	CAL2	1	Water	NA	NA	20.7	7.00	02/22/2023	16:21
3	CAL3	1	Water	NA	NA	20.7	10.03	02/22/2023	16:22
4	ICV	1	Water	NA	NA	20.7	6.99	02/22/2023	16:25
5	CCV1	1	Water	NA	NA	20.7	2.01	02/22/2023	16:30
6	O1634-01	1	Water	NA	NA	20.6	4.34	02/22/2023	16:33
7	O1634-01DUP	1	Water	NA	NA	20.7	4.35	02/22/2023	16:34
8	O1643-06	1	Water	NA	NA	20.6	7.79	02/22/2023	16:38
9	O1646-01	1	Water	NA	NA	20.5	4.24	02/22/2023	16:40
10	O1646-02	1	Water	NA	NA	20.5	4.39	02/22/2023	16:41
11	O1646-03	1	Water	NA	NA	20.2	4.43	02/22/2023	16:43
12	O1653-01	1	Water	NA	NA	20.4	11.72	02/22/2023	16:49
13	CCV2	1	Water	NA	NA	20.6	12.03	02/22/2023	16:50

18124202

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O1653 WorkList ID : 167604 Department : Wet-Chemistry Date : 02-22-2023 16:13:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/20/2023	Water	O1634-01	pH	Cool 4 deg C	STER12	K51	A-B-CELL-POOL-WATER-1ST-Q	02/20/2023	9040C
02/22/2023	Water	O1643-06	pH	Cool 4 deg C	PSEG03	D61	34292	02/22/2023	9040C
02/21/2023	Water	O1646-01	pH	Cool 4 deg C	STER13	D62	CELL-A-POOL-WATER	02/21/2023	9040C
02/21/2023	Water	O1646-02	pH	Cool 4 deg C	STER13	D62	CELL-B-POOL-WATER	02/21/2023	9040C
02/21/2023	Water	O1646-03	pH	Cool 4 deg C	STER13	D62	CELL-C-POOL-WATER	02/21/2023	9040C
02/22/2023	Water	O1653-01	pH	Cool 4 deg C	PSEG03	D62	OK-0567	02/22/2023	9040C

Date/Time 02/22/23 16:15

Raw Sample Received by: JO (W)

Raw Sample Relinquished by: JT (SM)

Date/Time 02/22/23 18:00

Raw Sample Received by: JT (SM)

Raw Sample Relinquished by: JO (W)