



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124203

Slope : 99.3

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	02/22/2023	16:20
2	CAL2	1	Water	NA	NA	20.7	7.00	02/22/2023	16:21
3	CAL3	1	Water	NA	NA	20.7	10.03	02/22/2023	16:22
4	ICV	1	Water	NA	NA	20.7	6.99	02/22/2023	16:25
5	CCV1	1	Water	NA	NA	20.7	2.01	02/22/2023	16:30
6	O1643-01	1	Solid	20.02	20	22.6	6.28	02/22/2023	17:00
7	O1643-02	1	Solid	20.03	20	23.3	6.28	02/22/2023	17:01
8	O1643-03	1	Solid	20.04	20	23.6	6.19	02/22/2023	17:05
9	O1643-04	1	Solid	20.02	20	23.4	4.59	02/22/2023	17:11
10	O1650-01	1	Solid	20.03	20	23.3	7.36	02/22/2023	17:15
11	O1650-01DUP	1	Solid	20.04	20	23.4	7.37	02/22/2023	17:16
12	CCV2	1	Water	NA	NA	20.7	12.03	02/22/2023	17:20

WORKLIST(Hardcopy Internal Chain)

18122203

WorkList Name : PH O1654 S

WorkList ID : 167605

Department : Wet-Chemistry

Date : 02-22-2023 16:14:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/22/2023	Solid	O1643-01	pH	Cool 4 deg C	PSEG03	D61	34143	02/22/2023	9045D
02/22/2023	Solid	O1643-02	pH	Cool 4 deg C	PSEG03	D61	34144	02/22/2023	9045D
02/22/2023	Solid	O1643-03	pH	Cool 4 deg C	PSEG03	D61	34145	02/22/2023	9045D
02/22/2023	Solid	O1643-04	pH	Cool 4 deg C	PSEG03	D61	34146	02/22/2023	9045D
02/22/2023	Solid	O1650-01	pH	Cool 4 deg C	PSEG05	D62	BU-2-022223	02/22/2023	9045D

Date/Time 02/22/23 16:20

Raw Sample Received by: 5012001

Raw Sample Relinquished by: JT(Sm)

Date/Time 02/22/23

Raw Sample Received by: JT(Sm)

Raw Sample Relinquished by: 2012001