


63124214

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 3.0264e-004

C= -1.2731e-002

Rho= 0.9979875

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	8	0.000	8				
0.2	0.200	0.172	-0.028	612	0.0 %	612				
2.5	2.500	2.441	-0.059	8109	0.0 %	8109				
5.0	5.000	5.005	0.005	16581	0.0 %	16581				
7.5	7.500	7.946	0.446	26296	0.0 %	26296				
10.0	10.000	9.645	-0.355	31913	0.0 %	31913				

LB124214 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Feb 2023 11:31:25

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	8	-	0.0000	7470A	PPB	23 Feb 2023 11:36:24	S	Std
0.2 - 1	50.2	612	-	0.2000	7470A	PPB	23 Feb 2023 11:38:41	S	Std
2.5 - 1	52.5	8109	-	2.5000	7470A	PPB	23 Feb 2023 11:40:59	S	Std
5.0 - 1	55	16581	-	5.0000	7470A	PPB	23 Feb 2023 11:45:49	S	Std
7.5 - 1	59.5	26296	-	7.5000	7470A	PPB	23 Feb 2023 11:48:06	S	Std
10.0 - 1	510	31913	-	10.0000	7470A	PPB	23 Feb 2023 11:52:58	S	Std
ICV59 - 1	ICV59	13889	4.1907	-	7470A	PPB	23 Feb 2023 11:58:20	U	SMPL
ICB59 - 1	ICB59	-29	-0.0215	-	7470A	PPB	23 Feb 2023 12:00:35	U	SMPL
CCV76 - 1	CCV76	16438	4.9621	-	7470A	PPB	23 Feb 2023 12:02:51	U	SMPL
CCB76 - 1	CCB76	-94	-0.0412	-	7470A	PPB	23 Feb 2023 12:05:06	U	SMPL
CRA - 1	CRA	600	0.1689	-	7470A	PPB	23 Feb 2023 12:07:21	U	SMPL
HighStd - 1	HighStd	32795	9.9124	-	7470A	PPB	23 Feb 2023 12:09:37	U	SMPL
ChkStd - 1	ChkStd	24539	7.4138	-	7470A	PPB	23 Feb 2023 12:11:53	U	SMPL
PB151015BL - 1	PBW	-80	-0.0369	-	7470A	PPB	23 Feb 2023 12:14:09	U	SMPL
PB151015BS - 1	LCSW	12570	3.7915	-	7470A	PPB	23 Feb 2023 12:16:25	U	SMPL
O1633-01 - 1	WEIR-TANK-MUD	-36	-0.0236	-	7470A	PPB	23 Feb 2023 12:18:41	U	SMPL
O1633-01DUP - 1	WEIR-TANK-MUDDUP	-47	-0.0270	-	7470A	PPB	23 Feb 2023 12:20:58	U	SMPL
O1633-01MS - 1	WEIR-TANK-MUDMS	12930	3.9004	-	7470A	PPB	23 Feb 2023 12:26:15	U	SMPL
O1633-01MSD - 1	WEIR-TANK-MUDMSD	12431	3.7494	-	7470A	PPB	23 Feb 2023 12:28:32	U	SMPL
O1638-02 - 1	VB-15-241	-30	-0.0218	-	7470A	PPB	23 Feb 2023 12:30:48	U	SMPL
CCV77 - 1	CCV77	16741	5.0538	-	7470A	PPB	23 Feb 2023 12:33:04	U	SMPL
CCB77 - 1	CCB77	-48	-0.0273	-	7470A	PPB	23 Feb 2023 12:35:19	U	SMPL
PB150998TB - 1	PB150998TB	-175	-0.0657	-	7470A	PPB	23 Feb 2023 13:13:49	U	SMPL
O1633-01LX5 - 1		-163	-0.0621	-	7470A	PPB	23 Feb 2023 13:16:07	U	SMPL
O1633-01A - 1		14891	4.4939	-	7470A	PPB	23 Feb 2023 13:18:24	U	SMPL
CCV79 - 1	CCV79	16504	4.9821	-	7470A	PPB	23 Feb 2023 13:20:41	U	SMPL
CCB79 - 1	CCB79	339	0.0899	-	7470A	PPB	23 Feb 2023 13:22:58	U	SMPL