



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB124223
Thermometer ID: Flashpoint

Reviewed By: Rubina
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 760.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	02/23/2023	10:00
2	O1643-06		1	12	100.00	>212.0	>212.0	02/23/2023	10:30
3	O1643-06DUP		1	12	100.00	>212.0	>212.0	02/23/2023	11:00
4	O1647-01		1	13	100.00	>212.0	>212.0	02/23/2023	11:30
5	O1653-01		1	14	100.00	>212.0	>212.0	02/23/2023	12:00
6	O1653-02		1	14	100.00	>212.0	>212.0	02/23/2023	12:30
7	O1653-03		1	14	100.00	>212.0	>212.0	02/23/2023	13:00
8	O1653-04		1	12	100.00	>212.0	>212.0	02/23/2023	13:30
9	O1653-05		1	14	100.00	>212.0	>212.0	02/23/2023	14:00
10	O1653-06		1	14	100.00	>212.0	>212.0	02/23/2023	14:30

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

66124223

WorkList Name: fp-2-22

WorkList ID: 167584

Department: Wet-Chemistry

Date: 02-22-2023 08:34:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/01/2023	Water	O1643-06	Flash Point	Cool 4 deg C	PSEG03	D61	34292	02/22/2023	1010B
02/27/2023	Water	O1647-01	Flash Point	Cool 4 deg C	PSEG03	D62	ACADEMY-SUBSTATION	02/22/2023	1010B
03/01/2023	Water	O1653-01	Flash Point	Cool 4 deg C	PSEG03	D62	OK-0567	02/22/2023	1010B
03/01/2023	Water	O1653-02	Flash Point	Cool 4 deg C	PSEG03	D62	31969	02/22/2023	1010B
03/01/2023	Water	O1653-03	Flash Point	Cool 4 deg C	PSEG03	D62	31967	02/22/2023	1010B
03/01/2023	Water	O1653-04	Flash Point	Cool 4 deg C	PSEG03	D62	31968	02/22/2023	1010B
03/01/2023	Water	O1653-05	Flash Point	Cool 4 deg C	PSEG03	D62	31966	02/22/2023	1010B
03/01/2023	Water	O1653-06	Flash Point	Cool 4 deg C	PSEG03	D62	OK-1555	02/22/2023	1010B

Date/Time 02/23/2023 09:00
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: AKCWC

Date/Time 02/23/2023 15:15
 Raw Sample Received by: AKCWC
 Raw Sample Relinquished by: RM CWC