



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124225

Slope : 98.8

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	02/23/2023	16:00
2	CAL2	1	Water	NA	NA	20.8	7.00	02/23/2023	16:02
3	CAL3	1	Water	NA	NA	20.7	10.03	02/23/2023	16:05
4	ICV	1	Water	NA	NA	20.7	6.99	02/23/2023	16:10
5	CCV1	1	Water	NA	NA	20.7	2.01	02/23/2023	16:11
6	O1654-02	1	Solid	20.02	20	22.8	8.32	02/23/2023	16:15
7	O1654-02DUP	1	Solid	20.03	20	22.7	8.33	02/23/2023	16:16
8	O1665-02	1	Solid	20.04	20	24.0	7.64	02/23/2023	16:20
9	CCV2	1	Water	NA	NA	20.8	12.03	02/23/2023	16:25

124225

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY O1665 WorkList ID : 167631 Department : Wet-Chemistry Date : 02-23-2023 13:28:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/22/2023	Solid	O1654-02	Corrosivity	Cool 4 deg C	PSEG03	D61	UGT	02/22/2023	9045D
02/23/2023	Solid	O1665-02	Corrosivity	Cool 4 deg C	PSEG03	E11	VNJ-231	02/23/2023	9045D

Date/Time 02-23-23 15:40
Raw Sample Received by: 70 (WJC)
Raw Sample Relinquished by: [Signature]

Date/Time 02-23-23 17:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: 70 (WJC)