

LB124238


SFAM01.1

INSTRUMENT 1A: CVI

Linear

 μ Abs.:

49117



A= 0.0000e+000

B= 2.0122e-004 slope

C= 5.2376e-003 y-intercept

Rho= 0.9996578

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o)
0.00	0.000	0.008	0.008	13	0.000	13					=
0.05	0.050					185	0				4
0.20	0.200	0.208	0.008	1010	0.0 %	1010					-5
2.50	2.500	2.368	-0.132	11743	0.0 %	11743					3
5.00	5.000	5.126	0.126	25450	0.0 %	25450					1
7.50	7.500	7.600	0.100	37744	0.0 %	37744					-1
10.0	10.000	9.889	-0.111	49117	0.0 %	49117					

LB124238

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	13	-		0 SFAM01.1	PPB	2/24/2023 14:59	Std
	0.2 S01	1010	-		0.2 SFAM01.1	PPB	2/24/2023 15:01	Std
	2.5 S02	11743	-		2.5 SFAM01.1	PPB	2/24/2023 15:04	Std
	5 S03	25450	-		5 SFAM01.1	PPB	2/24/2023 15:09	Std
	7.5 S04	37744	-		7.5 SFAM01.1	PPB	2/24/2023 15:11	Std
	10 S05	49117	-		10 SFAM01.1	PPB	2/24/2023 15:16	Std
ICV061	ICV061	19763	3.982 -		SFAM01.1	PPB	2/24/2023 15:23	SMPL
ICB061	ICB061	-46	-0.004 -		SFAM01.1	PPB	2/24/2023 15:25	SMPL
CCV080	CCV080	25284	5.093 -		SFAM01.1	PPB	2/24/2023 15:27	SMPL
CCB080	CCB080	-51	-0.005 -		SFAM01.1	PPB	2/24/2023 15:29	SMPL
PB151068BL	PBW068	11	0.0075 -		SFAM01.1	PPB	2/24/2023 15:32	SMPL
O1603-01	YBL49	599	0.1258 -		SFAM01.1	PPB	2/24/2023 15:34	SMPL
O1603-02	YBL49D	589	0.1238 -		SFAM01.1	PPB	2/24/2023 15:36	SMPL
O1603-03	YBL49S	6175	1.2478 -		SFAM01.1	PPB	2/24/2023 15:39	SMPL
PB151069BL	PBW069	-7	0.0038 -		SFAM01.1	PPB	2/24/2023 15:41	SMPL
O1604-01	YBL50	215	0.0485 -		SFAM01.1	PPB	2/24/2023 15:43	SMPL
O1604-02	YBL50D	219	0.0493 -		SFAM01.1	PPB	2/24/2023 15:45	SMPL
O1604-03	YBL50S	5714	1.155 -		SFAM01.1	PPB	2/24/2023 15:48	SMPL
O1604-04	YBL51	3	0.0058 -		SFAM01.1	PPB	2/24/2023 15:50	SMPL
CCV081	CCV081	25114	5.0588 -		SFAM01.1	PPB	2/24/2023 15:52	SMPL
CCB081	CCB081	-44	-0.0036 -		SFAM01.1	PPB	2/24/2023 15:54	SMPL