



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 3/8/2023

OVENTEMP IN Celsius(°C): 106
Time IN: 16:50
In Date: 02/24/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:33
Out Date: 02/25/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB124244

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1694-01	EXHP0	1	1.18	8.67	9.85	8.76	87.4	
O1694-02	EXHP1	2	1.13	8.78	9.91	8.85	87.9	
O1694-03	EXHP2	3	1.15	8.81	9.96	8.85	87.4	
O1694-04	EXHP3	4	1.16	8.60	9.76	8.36	83.7	
O1694-05	EXHP4	5	1.13	8.63	9.76	8.03	80.0	
O1694-06	EXHP5	6	1.15	8.81	9.96	8.75	86.3	
O1694-07	EXHP6	7	1.13	8.55	9.68	8.22	82.9	
O1694-08	EXHP7	8	1.19	8.56	9.75	8.04	80.0	
O1694-09	EXHP8	9	1.19	8.75	9.94	8.46	83.1	
O1694-10	EXHP9	10	1.19	8.78	9.97	8.08	78.5	
O1694-11	EXHQ0	11	1.13	8.53	9.66	7.97	80.2	
O1694-12	EXHQ1	12	1.17	8.66	9.83	8.34	82.8	
O1694-13	EXHQ2	13	1.11	8.44	9.55	8.07	82.5	
O1694-14	EXHQ3	14	1.15	8.79	9.94	8.64	85.2	
O1694-15	EXHQ4	15	1.13	8.60	9.73	7.86	78.3	
O1694-16	EXHQ5	16	1.17	8.49	9.66	8.28	83.7	
O1694-17	EXHQ6	17	1.18	8.54	9.72	8.31	83.5	
O1694-18	EXHQ7	18	1.15	8.83	9.98	8.44	82.6	
O1694-19	EXHQ8	19	1.1	8.61	9.71	8.22	82.7	
O1694-20	EXHQ9	20	1.18	8.76	9.94	8.05	78.4	
O1694-21	EXHQ9MS	21	1.18	8.76	9.94	8.05	78.4	
O1694-22	EXHQ9MSD	22	1.18	8.76	9.94	8.05	78.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$