

LB124250


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

63071



A= 0.0000e+000

B= 1.6022e-004

C= 4.5203e-002

Rho= 0.9996419

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.041	0.041	-29	0.000	-29					
0.2	0.200	0.232	0.032	1168	0.0 %	1168					16
2.5	2.500	2.521	0.021	15455	0.0 %	15455					1
5.0	5.000	4.915	-0.085	30394	0.0 %	30394					-2
7.5	7.500	7.340	-0.160	45532	0.0 %	45532					-2
10.0	10.000	10.150	0.150	63071	0.0 %	63071					2

LB124250 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Feb 2023 16:13:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-29	-	0.0000	7471B	PPB	24 Feb 2023 16:18:14	S	Std
0.2 - 1	50.2	1168	-	0.2000	7471B	PPB	24 Feb 2023 16:20:31	S	Std
2.5 - 1	52.5	15455	-	2.5000	7471B	PPB	24 Feb 2023 16:22:48	S	Std
5.0 - 1	55	30394	-	5.0000	7471B	PPB	24 Feb 2023 16:25:06	S	Std
7.5 - 1	57.5	45532	-	7.5000	7471B	PPB	24 Feb 2023 16:27:22	S	Std
10.0 - 1	510	63071	-	10.0000	7471B	PPB	24 Feb 2023 16:32:13	S	Std
ICV62 - 1	ICV62	25551	4.1390	-	7471B	PPB	24 Feb 2023 16:40:13	U	SMPL
ICB62 - 1	ICB62	-7	0.0441	-	7471B	PPB	24 Feb 2023 16:42:28	U	SMPL
CCV82 - 1	CCV82	30560	4.9415	-	7471B	PPB	24 Feb 2023 16:44:44	U	SMPL
CCB82 - 1	CCB82	-10	0.0436	-	7471B	PPB	24 Feb 2023 16:46:59	U	SMPL
CRA - 1	CRA	1136	0.2272	-	7471B	PPB	24 Feb 2023 16:49:16	U	SMPL
HighStd - 1	HighStd	63770	10.2624	-	7471B	PPB	24 Feb 2023 16:51:31	U	SMPL
ChkStd - 1	ChkStd	43205	6.9675	-	7471B	PPB	24 Feb 2023 16:53:47	U	SMPL
PB151051BL - 1	PBS	-111	0.0274	-	7471B	PPB	24 Feb 2023 16:56:05	U	SMPL
PB151051BS - 1	LCSS	22230	3.6069	-	7471B	PPB	24 Feb 2023 17:01:12	U	SMPL
O1645-01 - 1	ETGI-352	5969	1.0016	-	7471B	PPB	24 Feb 2023 17:03:28	U	SMPL
O1645-01DUP - 1	ETGI-352DUP	5993	1.0054	-	7471B	PPB	24 Feb 2023 17:09:23	U	SMPL
O1645-01MS - 1	ETGI-352MS	27866	4.5099	-	7471B	PPB	24 Feb 2023 17:11:39	U	SMPL
O1645-01MSD - 1	ETGI-352MSD	30830	4.9848	-	7471B	PPB	24 Feb 2023 17:13:56	U	SMPL
O1650-01 - 1	BU-2-022223	3509	0.6074	-	7471B	PPB	24 Feb 2023 17:16:12	U	SMPL
CCV83 - 1	CCV83	30353	4.9084	-	7471B	PPB	24 Feb 2023 17:18:29	U	SMPL
CCB83 - 1	CCB83	-89	0.0309	-	7471B	PPB	24 Feb 2023 17:20:44	U	SMPL
O1654-01 - 1	UGT	172113	27.6211	-	7471B	PPB	24 Feb 2023 17:23:00	U	SMPL
O1657-01 - 1	B-17-SB01	20336	3.3034	-	7471B	PPB	24 Feb 2023 17:25:15	U	SMPL
O1657-02 - 1	B-17-SB02	858	0.1827	-	7471B	PPB	24 Feb 2023 17:27:35	U	SMPL
O1658-01 - 1	QAQC-COMP-1	13922	2.2758	-	7471B	PPB	24 Feb 2023 17:29:51	U	SMPL
O1659-01 - 1	QAQC-COMP-2	23104	3.7469	-	7471B	PPB	24 Feb 2023 17:32:07	U	SMPL
O1660-01 - 1	QAQC-COMP-1	20366	3.3082	-	7471B	PPB	24 Feb 2023 17:34:23	U	SMPL
O1661-01 - 1	WC-11	9044	1.4942	-	7471B	PPB	24 Feb 2023 17:36:40	U	SMPL
O1662-01 - 1	WC-E2	10568	1.7384	-	7471B	PPB	24 Feb 2023 17:38:57	U	SMPL
O1663-01 - 1	WC-D2	15494	2.5277	-	7471B	PPB	24 Feb 2023 17:41:17	U	SMPL
O1665-01 - 1	VNJ-231	5504	0.9271	-	7471B	PPB	24 Feb 2023 17:43:33	U	SMPL
CCV84 - 1	CCV84	30338	4.9060	-	7471B	PPB	24 Feb 2023 17:45:49	U	SMPL
CCB84 - 1	CCB84	-68	0.0343	-	7471B	PPB	24 Feb 2023 17:48:04	U	SMPL
O1669-01 - 1	NB-08-022323	715	0.1598	-	7471B	PPB	24 Feb 2023 17:50:20	U	SMPL
O1671-01 - 1	T5-1	3394	0.5890	-	7471B	PPB	24 Feb 2023 17:52:36	U	SMPL
O1645-01LX5 - 1		883	0.1867	-	7471B	PPB	24 Feb 2023 17:54:51	U	SMPL
O1645-01A - 1		30775	4.9760	-	7471B	PPB	24 Feb 2023 17:57:07	U	SMPL
O1654-01DLX5 - 1		38358	6.1909	-	7471B	PPB	24 Feb 2023 17:59:23	U	SMPL
CCV85 - 1	CCV85	30398	4.9156	-	7471B	PPB	24 Feb 2023 18:01:40	U	SMPL
CCB85 - 1	CCB85	-69	0.0341	-	7471B	PPB	24 Feb 2023 18:03:58	U	SMPL