



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/28/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 02/27/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 02/28/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB124270

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1697-01	EXHM9	1	1.18	8.37	9.55	9.32	97.3	
O1697-02	EXHM9MS	2	1.18	8.37	9.55	9.32	97.3	
O1697-03	EXHM9MSD	3	1.18	8.37	9.55	9.32	97.3	
O1697-04	EXHX0	4	1.12	8.44	9.56	8.45	86.8	
O1697-05	EXHX1	5	1.19	8.58	9.77	8.79	88.6	
O1697-06	EXHX2	6	1.16	8.38	9.54	8.66	89.5	
O1697-07	EXHX3	7	1.19	8.43	9.62	8.55	87.3	
O1697-08	EXHX4	8	1.13	8.73	9.86	8.02	78.9	
O1697-09	EXHX5	9	1.18	8.79	9.97	8.53	83.6	
O1697-10	EXHX6	10	1.1	8.69	9.79	8.17	81.4	
O1697-11	EYF98	11	1.13	8.67	9.8	8.09	80.3	
O1697-12	EYF99	12	1.15	8.53	9.68	7.92	79.4	
O1697-13	EYFA0	13	1.19	8.59	9.78	7.91	78.2	
O1697-14	EYFA1	14	1.19	8.50	9.69	8.34	84.1	
O1697-15	EYFA2	15	1.15	8.79	9.94	8.59	84.6	
O1697-16	EYFA3	16	1.19	8.79	9.98	9.4	93.4	
O1697-17	EYFA4	17	1.18	8.78	9.96	8.41	82.3	
O1697-18	EYFA5	18	1.16	8.58	9.74	8.2	82.1	
O1697-19	EYFA6	19	1.13	8.54	9.67	8.56	87.0	
O1697-20	EYFA7	20	1.18	8.37	9.55	8.51	87.6	
O1697-21	EYFA8	21	1.19	8.69	9.88	9.05	90.4	
O1697-22	EYFA9	22	1.18	8.44	9.62	8.69	89.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$