



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Rubina

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124272

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	02/27/2023	11:25
2	CAL2	1	Water	NA	NA	20.7	7.01	02/27/2023	11:26
3	CAL3	1	Water	NA	NA	20.8	10.03	02/27/2023	11:27
4	ICV	1	Water	NA	NA	20.7	6.99	02/27/2023	11:30
5	CCV1	1	Water	NA	NA	20.8	2.01	02/27/2023	11:32
6	O1647-01	1	Water	NA	NA	19.5	8.62	02/27/2023	11:35
7	O1647-01DUP	1	Water	NA	NA	19.5	8.61	02/27/2023	11:38
8	CCV2	1	Water	NA	NA	20.7	12.01	02/27/2023	11:40

WORKLIST(Hardcopy Internal Chain)

LB124272

WorkList Name : pH-022723 WorkList ID : 167711 Department : Wet-Chemistry Date : 02-27-2023 11:15:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/22/2023	Water	O1647-01	pH	Cool 4 deg C	PSEG03	--Sele	ACADEMY-SUBSTATION	02/22/2023	9040C

Date/Time 02/27/23 11:20
Raw Sample Received by: RM (CWC)
Raw Sample Relinquished by: RM (CWC)

Date/Time 02/27/23 11:45
Raw Sample Received by: RM (CWC)
Raw Sample Relinquished by: TPCWCY