

LB124273


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

45372



A= 0.0000e+000

B= 2.1987e-004

C= 1.9546e-002

Rho= 0.9999007

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.025	0.025	23	0.000	23					
0.2	0.200	0.188	-0.012	768	0.0 %	768					1
2.5	2.500	2.532	0.032	11429	0.0 %	11429					-6
5.0	5.000	4.897	-0.103	22184	0.0 %	22184					-2
7.5	7.500	7.562	0.062	34304	0.0 %	34304					1
10.0	10.000	9.995	-0.005	45372	0.0 %	45372					0

OLB124273 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Feb 2023 12:18:07

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7470A	PPB	27 Feb 2023 12:38:31	S	Std
0.2 - 1	50.2	768	-	0.2000	7470A	PPB	27 Feb 2023 12:40:48	S	Std
2.5 - 1	52.5	11429	-	2.5000	7470A	PPB	27 Feb 2023 12:43:06	S	Std
5.0 - 1	55	22184	-	5.0000	7470A	PPB	27 Feb 2023 12:45:23	S	Std
7.5 - 1	57.5	34304	-	7.5000	7470A	PPB	27 Feb 2023 12:47:39	S	Std
10.0 - 1	59	45372	-	10.0000	7470A	PPB	27 Feb 2023 12:52:42	S	Std
ICV63 - 1	ICV63	19655	4.3411	-	7470A	PPB	27 Feb 2023 12:55:29	U	SMPL
ICB63 - 1	ICB63	-70	0.0042	-	7470A	PPB	27 Feb 2023 12:57:44	U	SMPL
CCV86 - 1	CCV86	22533	4.9739	-	7470A	PPB	27 Feb 2023 12:59:59	U	SMPL
CCB86 - 1	CCB86	-73	0.0035	-	7470A	PPB	27 Feb 2023 13:02:14	U	SMPL
CRA - 1	CRA	807	0.1970	-	7470A	PPB	27 Feb 2023 13:04:29	U	SMPL
HighStd - 1	HighStd	44553	9.8154	-	7470A	PPB	27 Feb 2023 13:06:45	U	SMPL
ChkStd - 1	ChkStd	31344	6.9111	-	7470A	PPB	27 Feb 2023 13:09:00	U	SMPL
PB151078BL - 1	PBW	-89	-0.0000	-	7470A	PPB	27 Feb 2023 13:11:17	U	SMPL
PB151078BS - 1	LCSW	16980	3.7529	-	7470A	PPB	27 Feb 2023 13:13:34	U	SMPL
O1654-02 - 1	UGT	479	0.1249	-	7470A	PPB	27 Feb 2023 13:15:50	U	SMPL
O1665-02 - 1	VNJ-231	431	0.1143	-	7470A	PPB	27 Feb 2023 13:18:07	U	SMPL
O1665-02DUP - 1	VNJ-231DUP	405	0.1086	-	7470A	PPB	27 Feb 2023 13:20:24	U	SMPL
O1665-02MS - 1	VNJ-231MS	18262	4.0348	-	7470A	PPB	27 Feb 2023 13:22:41	U	SMPL
O1665-02MSD - 1	VNJ-231MSD	17257	3.8138	-	7470A	PPB	27 Feb 2023 13:24:57	U	SMPL
CCV87 - 1	CCV87	22668	5.0036	-	7470A	PPB	27 Feb 2023 13:27:12	U	SMPL
CCB87 - 1	CCB87	-85	0.0053	-	7470A	PPB	27 Feb 2023 13:29:28	U	SMPL
PB151079BL - 1	PBW	-21	0.0149	-	7470A	PPB	27 Feb 2023 13:31:44	U	SMPL
PB151079BS - 1	LCSW	18453	4.0768	-	7470A	PPB	27 Feb 2023 13:33:59	U	SMPL
O1647-01 - 1	ACADEMY-SUBSTATION	480	0.1251	-	7470A	PPB	27 Feb 2023 13:36:15	U	SMPL
O1647-01DUP - 1	ACADEMY-SUBSTATIONDUP	517	0.1332	-	7470A	PPB	27 Feb 2023 13:38:30	U	SMPL
O1647-01MS - 1	ACADEMY-SUBSTATIONMS	18675	4.1256	-	7470A	PPB	27 Feb 2023 13:40:46	U	SMPL
O1647-01MSD - 1	ACADEMY-SUBSTATIONMSD	18645	4.1190	-	7470A	PPB	27 Feb 2023 13:43:03	U	SMPL
PB151037BL - 1	PBW	-80	0.0020	-	7470A	PPB	27 Feb 2023 13:45:20	U	SMPL
PB151037BS - 1	LCSW	17629	3.8956	-	7470A	PPB	27 Feb 2023 13:47:37	U	SMPL
O1641-01 - 1	B-26-GW01	1167	0.2761	-	7470A	PPB	27 Feb 2023 13:49:54	U	SMPL
O1641-01DUP - 1	B-26-GW01DUP	1101	0.2616	-	7470A	PPB	27 Feb 2023 13:52:11	U	SMPL
CCV88 - 1	CCV88	22614	4.9917	-	7470A	PPB	27 Feb 2023 13:54:26	U	SMPL
CCB88 - 1	CCB88	-98	-0.0020	-	7470A	PPB	27 Feb 2023 13:56:42	U	SMPL
O1641-01MS - 1	B-26-GW01MS	18795	4.1520	-	7470A	PPB	27 Feb 2023 13:58:58	U	SMPL
O1641-01MSD - 1	B-26-GW01MSD	18970	4.1905	-	7470A	PPB	27 Feb 2023 14:01:13	U	SMPL
O1653-01DLX100 - 1	OK-0567	-80	0.0020	-	7470A	PPB	27 Feb 2023 14:03:29	U	SMPL
O1653-02 - 1	31969	122	0.0464	-	7470A	PPB	27 Feb 2023 14:05:45	U	SMPL
O1653-03 - 1	31967	268	0.0780	-	7470A	PPB	27 Feb 2023 14:08:01	U	SMPL
O1653-04 - 1	31968	10211	2.2646	-	7470A	PPB	27 Feb 2023 14:10:18	U	SMPL
O1653-05 - 1	31966	208	0.0653	-	7470A	PPB	27 Feb 2023 14:12:35	U	SMPL
O1653-06DLX10 - 1	OK-1555	25474	5.6205	-	7470A	PPB	27 Feb 2023 14:14:52	U	SMPL
O1655-10 - 1	31994	131	0.0483	-	7470A	PPB	27 Feb 2023 14:17:09	U	SMPL
O1657-03 - 1	EB-2023-2-22	-42	0.0103	-	7470A	PPB	27 Feb 2023 14:19:25	U	SMPL
CCV89 - 1	CCV89	22794	5.0313	-	7470A	PPB	27 Feb 2023 14:21:41	U	SMPL
CCB89 - 1	CCB89	-49	0.0088	-	7470A	PPB	27 Feb 2023 14:23:56	U	SMPL
O1701-03 - 1	B-13-GW01	1853	0.4270	-	7470A	PPB	27 Feb 2023 14:26:12	U	SMPL
O1701-04 - 1	B-13-GW02	7215	1.6059	-	7470A	PPB	27 Feb 2023 14:28:28	U	SMPL
O1701-05 - 1	EB-2023-2-24	222	0.0684	-	7470A	PPB	27 Feb 2023 14:30:44	U	SMPL
PB151039TB - 1	PB151039TB	133	0.0488	-	7470A	PPB	27 Feb 2023 14:33:00	U	SMPL
PB151046TB - 1	PB151046TB	41	0.0286	-	7470A	PPB	27 Feb 2023 14:35:16	U	SMPL
CCV90 - 1	CCV90	21745	4.8006	-	7470A	PPB	27 Feb 2023 14:37:32	U	SMPL
CCB90 - 1	CCB90	-73	0.0035	-	7470A	PPB	27 Feb 2023 14:39:48	U	SMPL
O1665-02LX5 - 1		50	0.0305	-	7470A	PPB	27 Feb 2023 14:42:04	U	SMPL
O1665-02A - 1		19827	4.3789	-	7470A	PPB	27 Feb 2023 14:44:20	U	SMPL
O1647-01LX5 - 1		23	0.0246	-	7470A	PPB	27 Feb 2023 14:46:36	U	SMPL
O1647-01A - 1		18823	4.1582	-	7470A	PPB	27 Feb 2023 14:48:52	U	SMPL
O1641-01LX5 - 1		138	0.0499	-	7470A	PPB	27 Feb 2023 14:51:08	U	SMPL
O1641-01A - 1		19936	4.4029	-	7470A	PPB	27 Feb 2023 14:53:24	U	SMPL
CCV91 - 1	CCV91	21918	4.8386	-	7470A	PPB	27 Feb 2023 14:55:40	U	SMPL
CCB91 - 1	CCB91	-63	0.0057	-	7470A	PPB	27 Feb 2023 14:57:56	U	SMPL