

LB124304


7471B

INSTRUMENT ID: (V)

Linear

μ Abs.:

58863



A= 0.0000e+000

B= 1.6763e-004

C= 2.8488e-003

Rho= 0.9997120

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.008	-0.008	-63	0.000	-63					
0.2	0.200	0.174	-0.026	1019	0.0 %	1019					-13
2.5	2.500	2.497	-0.003	14878	0.0 %	14878					0
5.0	5.000	4.995	-0.005	29782	0.0 %	29782					0
7.5	7.500	7.672	0.172	45747	0.0 %	45747					2
10.0	10.000	9.870	-0.130	58863	0.0 %	58863					1

LB124304 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Feb 2023 13:34:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-63	-	0.0000	7471B	PPB	28 Feb 2023 13:41:53	S	Std
0.2 - 1	50.2	1019	-	0.2000	7471B	PPB	28 Feb 2023 13:44:10	S	Std
2.5 - 1	52.5	14878	-	2.5000	7471B	PPB	28 Feb 2023 13:46:28	S	Std
5.0 - 1	55	29782	-	5.0000	7471B	PPB	28 Feb 2023 13:48:45	S	Std
7.5 - 1	57.5	45747	-	7.5000	7471B	PPB	28 Feb 2023 13:51:01	S	Std
10.0 - 1	510	58863	-	10.0000	7471B	PPB	28 Feb 2023 13:53:18	S	Std
ICV64 - 1	ICV64	24149	4.0510	-	7471B	PPB	28 Feb 2023 13:58:35	U	SMPL
ICB64 - 1	ICB64	-61	-0.0074	-	7471B	PPB	28 Feb 2023 14:00:50	U	SMPL
CCV92 - 1	CCV92	29982	5.0288	-	7471B	PPB	28 Feb 2023 14:03:06	U	SMPL
CCB92 - 1	CCB92	-75	-0.0097	-	7471B	PPB	28 Feb 2023 14:05:21	U	SMPL
CRA - 1	CRA	1198	0.2037	-	7471B	PPB	28 Feb 2023 14:07:38	U	SMPL
HighStd - 1	HighStd	59659	10.0037	-	7471B	PPB	28 Feb 2023 14:09:53	U	SMPL
ChkStd - 1	ChkStd	42900	7.1943	-	7471B	PPB	28 Feb 2023 14:12:09	U	SMPL
PB151109BL - 1	PBS	-128	-0.0186	-	7471B	PPB	28 Feb 2023 14:14:26	U	SMPL
PB151109BS - 1	LCSS	21475	3.6028	-	7471B	PPB	28 Feb 2023 14:18:40	U	SMPL
O1692-01 - 1	FM1282	6378	1.0720	-	7471B	PPB	28 Feb 2023 14:20:56	U	SMPL
O1700-01 - 1	MW-1R	5852	0.9838	-	7471B	PPB	28 Feb 2023 14:23:13	U	SMPL
O1701-01 - 1	B-19-SB01	7171	1.2049	-	7471B	PPB	28 Feb 2023 14:25:30	U	SMPL
O1701-02 - 1	B-19-SB02	1296	0.2201	-	7471B	PPB	28 Feb 2023 14:27:47	U	SMPL
O1703-05 - 1	T2-STONE	140	0.0263	-	7471B	PPB	28 Feb 2023 14:30:03	U	SMPL
CCV93 - 1	CCV93	30033	5.0374	-	7471B	PPB	28 Feb 2023 14:32:18	U	SMPL
CCB93 - 1	CCB93	-69	-0.0087	-	7471B	PPB	28 Feb 2023 14:34:34	U	SMPL
O1722-01 - 1	HD-01-022723	828	0.1416	-	7471B	PPB	28 Feb 2023 14:36:51	U	SMPL
O1724-01 - 1	RBR53-VNJ201	13108	2.2002	-	7471B	PPB	28 Feb 2023 14:39:06	U	SMPL
O1725-01 - 1	EO-01-022723	3053	0.5146	-	7471B	PPB	28 Feb 2023 14:41:22	U	SMPL
O1726-01 - 1	SU-03-022723	2821	0.4757	-	7471B	PPB	28 Feb 2023 14:43:37	U	SMPL
O1727-01 - 1	JB-1S	308	0.0545	-	7471B	PPB	28 Feb 2023 14:45:53	U	SMPL
O1728-01 - 1	OR-03-022723	2407	0.4063	-	7471B	PPB	28 Feb 2023 14:48:10	U	SMPL
O1728-01DUP - 1	OR-03-022723DUP	2590	0.4370	-	7471B	PPB	28 Feb 2023 14:50:27	U	SMPL
O1728-01MS - 1	OR-03-022723MS	24223	4.0634	-	7471B	PPB	28 Feb 2023 14:52:44	U	SMPL
O1728-01MSD - 1	OR-03-022723MSD	25243	4.2344	-	7471B	PPB	28 Feb 2023 14:55:01	U	SMPL
CCV94 - 1	CCV94	29938	5.0215	-	7471B	PPB	28 Feb 2023 14:57:17	U	SMPL
CCB94 - 1	CCB94	-79	-0.0104	-	7471B	PPB	28 Feb 2023 14:59:33	U	SMPL
O1728-01LX5 - 1		323	0.0570	-	7471B	PPB	28 Feb 2023 15:01:50	U	SMPL
O1728-01A - 1		27029	4.5338	-	7471B	PPB	28 Feb 2023 15:04:05	U	SMPL
CCV95 - 1	CCV95	30249	5.0736	-	7471B	PPB	28 Feb 2023 15:06:21	U	SMPL
CCB95 - 1	CCB95	-58	-0.0069	-	7471B	PPB	28 Feb 2023 15:08:38	U	SMPL