

CB124310


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

41101



A= 0.0000e+000

B= 2.4062e-004

C= 1.8724e-002

Rho= 0.9994163

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.024	0.024	24	0.000	24					
0.2	0.200	0.211	0.011	797	0.0 %	797					
2.5	2.500	2.310	-0.190	9522	0.0 %	9522					
5.0	5.000	5.223	0.223	21627	0.0 %	21627					
7.5	7.500	7.524	0.024	31190	0.0 %	31190					
10.0	10.000	9.909	-0.091	41101	0.0 %	41101					

0.00
1.5
-0.8
4
0
-1

LB124310 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Feb 2023 15:51:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	56	24	-	0.0000	7470A	PPB	28 Feb 2023 15:59:30	S	Std
0.2 - 1	50.2	797	-	0.2000	7470A	PPB	28 Feb 2023 16:01:47	S	Std
2.5 - 1	52.1	9522	-	2.5000	7470A	PPB	28 Feb 2023 16:04:04	S	Std
5.0 - 1	55	21627	-	5.0000	7470A	PPB	28 Feb 2023 16:06:21	S	Std
7.5 - 1	57.5	31190	-	7.5000	7470A	PPB	28 Feb 2023 16:08:38	S	Std
10.0 - 1	510	41101	-	10.0000	7470A	PPB	28 Feb 2023 16:10:54	S	Std
ICV65 - 1	ICV65	17292	4.1796	-	7470A	PPB	28 Feb 2023 16:16:08	U	SMPL
ICB65 - 1	ICB65	-34	0.0105	-	7470A	PPB	28 Feb 2023 16:18:23	U	SMPL
CCV96 - 1	CCV96	21760	5.2547	-	7470A	PPB	28 Feb 2023 16:20:38	U	SMPL
CCB96 - 1	CCB96	-43	0.0084	-	7470A	PPB	28 Feb 2023 16:22:53	U	SMPL
CRA - 1	CRA	780	0.2064	-	7470A	PPB	28 Feb 2023 16:25:09	U	SMPL
HighStd - 1	HighStd	41931	10.1083	-	7470A	PPB	28 Feb 2023 16:27:24	U	SMPL
ChkStd - 1	ChkStd	29021	7.0019	-	7470A	PPB	28 Feb 2023 16:29:39	U	SMPL
PB151110BL - 1	PBW	-54	0.0057	-	7470A	PPB	28 Feb 2023 16:31:56	U	SMPL
PB151110BS - 1	LCSW	17175	4.1514	-	7470A	PPB	28 Feb 2023 16:34:12	U	SMPL
O1691-01 - 1	72-11995	-12	0.0158	-	7470A	PPB	28 Feb 2023 16:36:28	U	SMPL
O1691-02 - 1	VNJ-229	22	0.0240	-	7470A	PPB	28 Feb 2023 16:38:45	U	SMPL
O1692-02 - 1	FM1282	50	0.0308	-	7470A	PPB	28 Feb 2023 16:41:02	U	SMPL
O1700-02 - 1	MW-1R	31	0.0262	-	7470A	PPB	28 Feb 2023 16:43:19	U	SMPL
O1703-06 - 1	T2-STONE	21	0.0238	-	7470A	PPB	28 Feb 2023 16:45:34	U	SMPL
CCV97 - 1	CCV97	22567	5.4489	-	7470A	PPB	28 Feb 2023 16:47:50	U	SMPL
CCB97 - 1	CCB97	-38	0.0096	-	7470A	PPB	28 Feb 2023 16:50:06	U	SMPL
O1724-02 - 1	RBR53-VNJ201	19	0.0233	-	7470A	PPB	28 Feb 2023 16:52:21	U	SMPL
O1727-02 - 1	JB-1S	33	0.0267	-	7470A	PPB	28 Feb 2023 16:54:37	U	SMPL
O1727-02DUP - 1	JB-1SDUP	16	0.0226	-	7470A	PPB	28 Feb 2023 16:56:54	U	SMPL
O1727-02MS - 1	JB-1SMS	18463	4.4614	-	7470A	PPB	28 Feb 2023 16:59:11	U	SMPL
O1727-02MSD - 1	JB-1SMSD	17903	4.3266	-	7470A	PPB	28 Feb 2023 17:01:29	U	SMPL
PB151121BL - 1	PBW	-38	0.0096	-	7470A	PPB	28 Feb 2023 17:03:46	U	SMPL
PB151121BS - 1	LCSW	16824	4.0670	-	7470A	PPB	28 Feb 2023 17:20:51	U	SMPL
O1727-04 - 1	JB1-WATER	119	0.0474	-	7470A	PPB	28 Feb 2023 17:23:08	U	SMPL
O1727-04DUP - 1	JB1-WATERDUP	71	0.0358	-	7470A	PPB	28 Feb 2023 17:25:26	U	SMPL
O1727-04MS - 1	JB1-WATERMS	16429	3.9719	-	7470A	PPB	28 Feb 2023 17:27:42	U	SMPL
CCV98 - 1	CCV98	22618	5.4611	-	7470A	PPB	28 Feb 2023 17:29:58	U	SMPL
CCB98 - 1	CCB98	-66	0.0028	-	7470A	PPB	28 Feb 2023 17:32:13	U	SMPL
O1727-04MSD - 1	JB1-WATERMSD	17489	4.2270	-	7470A	PPB	28 Feb 2023 17:34:30	U	SMPL
PB151080TB - 1	PB151080TB	6	0.0202	-	7470A	PPB	28 Feb 2023 17:36:45	U	SMPL
O1727-02LX5 - 1		-17	0.0146	-	7470A	PPB	28 Feb 2023 17:39:02	U	SMPL
O1727-02A - 1		18066	4.3658	-	7470A	PPB	28 Feb 2023 17:41:19	U	SMPL
O1727-04LX5 - 1		-29	0.0117	-	7470A	PPB	28 Feb 2023 17:43:37	U	SMPL
O1727-04A - 1		18330	4.4294	-	7470A	PPB	28 Feb 2023 17:45:53	U	SMPL
CCV99 - 1	CCV99	21392	5.1661	-	7470A	PPB	28 Feb 2023 17:48:10	U	SMPL
CCB99 - 1	CCB99	-65	0.0031	-	7470A	PPB	28 Feb 2023 17:50:27	U	SMPL