

Analytical Summary Report

Analysis Method: SM5220 D
 Parameter: COD
 Run Number: LB124313

ANALYST: Iwona
 SUPERVISOR REVIEW BY: Sohil

Reagent/Standard	Lot/Log #
COD calibration std. 10 ppm	WP100109
COD calibration std. 150 ppm	WP100112
COD calibration std. 50 ppm	WP100110
COD calibration std. 0 ppm	WP100108
COD ICV-LCS std, 50ppm	WP100114
COD calibration std. 100 ppm	WP100111
COD Digestion Vials Low Level 0-150Mg/L	W2997
COD ICV-LCS std, 50ppm	WP101124
COD CCV std, 50ppm	WP101123
COD LOD std, 5ppm	WP101125

Temp In (C): <u>148</u>	Date In: <u>02/28/2023</u>	Time In: <u>10:20</u>
Temp Out (C): <u>150</u>	Date Out: <u>02/28/2023</u>	Time Out: <u>12:20</u>

Intercept: -0.039 Slope: 1.0006 Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	0.039		12/05/2022	15:25
2	CAL2	10	1	Water	11.000	11.032	10.3	12/05/2022	15:25
3	CAL3	50	1	Water	49.000	49.01	-2	12/05/2022	15:26
4	CAL4	100	1	Water	99.000	98.98	-1	12/05/2022	15:26
5	CAL5	150	1	Water	151.000	150.948	0.6	12/05/2022	15:27



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB124313

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	51.000	51.008	12/05/2022	15:27
2	ICB		NA	NA	1	Water	0.000	0.039	12/05/2022	15:28
3	CCV1	50	NA	NA	1	Water	50.000	50.009	02/28/2023	15:50
4	CCB1		NA	NA	1	Water	1.000	1.038	02/28/2023	15:50
5	LB124313BL		1	50	1	Solid	1.000	1.038	02/28/2023	15:51
6	LB124313BS	50	1	50	1	Solid	50.000	50.009	02/28/2023	15:51
7	O1627-03		1	50	1	Solid	5.000	5.036	02/28/2023	15:52
8	CCV2	50	NA	NA	1	Water	50.000	50.009	02/28/2023	15:52
9	CCB2		NA	NA	1	Water	1.000	1.038	02/28/2023	15:53