



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124316

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/01/2023	07:00
2	CAL2	1	Water	NA	NA	20.7	7.00	03/01/2023	07:02
3	CAL3	1	Water	NA	NA	20.7	10.03	03/01/2023	07:05
4	ICV	1	Water	NA	NA	20.8	6.98	03/01/2023	07:10
5	CCV1	1	Water	NA	NA	20.8	2.01	03/01/2023	07:11
6	O1724-02	1	Solid	20.02	20	23.2	8.95	03/01/2023	07:15
7	O1727-02	1	Solid	20.03	20	21.9	7.82	03/01/2023	07:20
8	O1739-02	1	Solid	20.02	20	21.6	8.69	03/01/2023	07:26
9	O1739-09	1	Solid	20.03	20	22.1	8.63	03/01/2023	07:30
10	O1739-16	1	Solid	20.02	20	21.0	8.82	03/01/2023	07:33
11	O1740-02	1	Solid	20.03	20	22.2	9.26	03/01/2023	07:35
12	O1740-02DUP	1	Solid	20.04	20	22.3	9.27	03/01/2023	07:36
13	CCV2	1	Water	NA	NA	20.8	12.03	03/01/2023	07:40

VB 124316

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY O1740

WorkList ID : 167756

Department : Wet-Chemistry

Date : 03-01-2023 06:07:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/27/2023	Solid	O1724-02	Corrosivity	Cool 4 deg C	PSEG03	E12	RBR53-VNJ201	02/27/2023	9045D
02/27/2023	Solid	O1727-02	Corrosivity	Cool 4 deg C	PSEG03	E12	JB-1S	02/27/2023	9045D
02/28/2023	Solid	O1739-02	Corrosivity	Cool 4 deg C	SOUN03	E21	SAMPLE-1-COMP-1	02/28/2023	9045D
02/28/2023	Solid	O1739-09	Corrosivity	Cool 4 deg C	SOUN03	E21	SAMPLE-2-COMP-2	02/28/2023	9045D
02/28/2023	Solid	O1739-16	Corrosivity	Cool 4 deg C	SOUN03	E21	SAMPLE-3-COMP-3	02/28/2023	9045D
02/28/2023	Solid	O1740-02	Corrosivity	Cool 4 deg C	SOUN03	E21	SAMPLE-4-COMP-4	02/28/2023	9045D

Date/Time 03-01-23 06:15
Raw Sample Received by: JB (WC)
Raw Sample Relinquished by: JB (WC)

Date/Time 03-01-23 9:00
Raw Sample Received by: JB (WC)
Raw Sample Relinquished by: JB (WC)