



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 3/2/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:10
In Date: 03/01/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 03/02/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB124319

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1711-01	DBAB7	1	1.13	8.60	9.73	8.61	87.0	
O1711-02	DBAB8	2	1.13	8.65	9.78	8.32	83.1	
O1711-03	DBAB9	3	1.13	8.69	9.82	7.92	78.1	
O1711-04	DBAC0	4	1.13	8.50	9.63	7.64	76.6	
O1711-05	DBAC1	5	1.12	8.85	9.97	7.28	69.6	
O1711-06	DBAC2	6	1.12	8.48	9.6	7.73	77.9	
O1711-07	DBAC3	7	1.13	8.40	9.53	7.41	74.8	
O1711-08	DBAC4	8	1.13	8.68	9.81	8.23	81.8	
O1711-09	DBAC5	9	1.13	8.58	9.71	7.74	77.0	
O1711-10	DBAC6	10	1.14	8.61	9.75	8.39	84.2	
O1711-11	DBAC7	11	1.13	8.85	9.98	8.92	88.0	
O1711-12	DBAC8	12	1.12	8.79	9.91	8.2	80.5	
O1711-13	DBAC9	13	1.13	8.42	9.55	7.3	73.3	
O1711-14	DBAD0	14	1.14	8.38	9.52	7.63	77.4	
O1711-15	DBAD0MS	15	1.14	8.38	9.52	7.63	77.4	
O1711-16	DBAD0MSD	16	1.14	8.38	9.52	7.63	77.4	
O1711-17	DBAD1	17	1.14	8.45	9.59	8.13	82.7	
O1711-18	DBAD2	18	1.14	8.63	9.77	8.05	80.1	
O1711-19	DBAD3	19	1.12	8.86	9.98	8.77	86.3	
O1711-20	DBAD4	20	1.15	8.83	9.98	8.5	83.2	
O1711-21	DBAD5	21	1.15	8.83	9.98	6.15	56.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$