


LB124323

7470A

INSTRUMENT ID: CVL

Linear



A= 0.0000e+000

B= 2.4352e-004

C= -4.5265e-002

Rho= 0.9996308

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.018	-0.018	110	0.000	110					100
0.2	0.200	0.252	0.052	1220	0.0 %	1220					26
2.5	2.500	2.345	-0.155	9817	0.0 %	9817					6
5.0	5.000	5.166	0.166	21399	0.0 %	21399					3
7.5	7.500	7.527	0.027	31096	0.0 %	31096					0
10.0	10.000	9.928	-0.072	40956	0.0 %	40956					1

LB124323 IMNSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Mar 2023 13:48:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	110	-	0.0000	7470A	PPB	01 Mar 2023 13:51:51	S	Std
0.2 - 1	50.2	1220	-	0.2000	7470A	PPB	01 Mar 2023 13:54:08	S	Std
2.5 - 1	52.5	9817	-	2.5000	7470A	PPB	01 Mar 2023 13:56:25	S	Std
5.0 - 1	55	21399	-	5.0000	7470A	PPB	01 Mar 2023 13:58:45	S	Std
7.5 - 1	7.5	31096	-	7.5000	7470A	PPB	01 Mar 2023 14:01:01	S	Std
10.0 - 1	510	40956	-	10.0000	7470A	PPB	01 Mar 2023 14:05:52	S	Std
ICV66 - 1	ICV66	17595	4.2395	-	7470A	PPB	01 Mar 2023 14:08:36	U	SMPL
ICB66 - 1	ICB66	489	0.0738	-	7470A	PPB	01 Mar 2023 14:10:51	U	SMPL
CCV01 - 1	CCV01	21361	5.1565	-	7470A	PPB	01 Mar 2023 14:13:06	U	SMPL
CCB01 - 1	CCB01	503	0.0772	-	7470A	PPB	01 Mar 2023 14:15:21	U	SMPL
CRA - 1	CRA	1199	0.2467	-	7470A	PPB	01 Mar 2023 14:17:37	U	SMPL
HighStd - 1	HighStd	40674	9.8596	-	7470A	PPB	01 Mar 2023 14:19:52	U	SMPL
ChkStd - 1	ChkStd	29031	7.0243	-	7470A	PPB	01 Mar 2023 14:22:08	U	SMPL
PB151143BL - 1	PBW	290	0.0254	-	7470A	PPB	01 Mar 2023 14:24:25	U	SMPL
PB151143BS - 1	LCSW	16193	3.8980	-	7470A	PPB	01 Mar 2023 14:26:42	U	SMPL
O1739-02 - 1	SAMPLE-1-COMP-1	448	0.0638	-	7470A	PPB	01 Mar 2023 14:28:58	U	SMPL
O1739-09 - 1	SAMPLE-2-COMP-2	474	0.0702	-	7470A	PPB	01 Mar 2023 14:31:15	U	SMPL
O1739-16 - 1	SAMPLE-3-COMP-3	487	0.0733	-	7470A	PPB	01 Mar 2023 14:33:32	U	SMPL
O1740-02 - 1	SAMPLE-4-COMP-4	489	0.0738	-	7470A	PPB	01 Mar 2023 14:35:49	U	SMPL
O1740-02DUP - 1	SAMPLE-4-COMP-4DUP	471	0.0694	-	7470A	PPB	01 Mar 2023 14:38:05	U	SMPL
CCV02 - 1	CCV02	21571	5.2077	-	7470A	PPB	01 Mar 2023 14:40:20	U	SMPL
CCB02 - 1	CCB02	526	0.0828	-	7470A	PPB	01 Mar 2023 14:42:36	U	SMPL
O1740-02MS - 1	SAMPLE-4-COMP-4MS	16441	3.9584	-	7470A	PPB	01 Mar 2023 14:44:51	U	SMPL
O1740-02MSD - 1	SAMPLE-4-COMP-4MSD	16296	3.9231	-	7470A	PPB	01 Mar 2023 14:47:07	U	SMPL
PB151111TB - 1	PB151111TB	435	0.0607	-	7470A	PPB	01 Mar 2023 14:49:22	U	SMPL
O1740-02LX5 - 1		64	-0.0297	-	7470A	PPB	01 Mar 2023 14:51:38	U	SMPL
O1740-02A - 1		16936	4.0790	-	7470A	PPB	01 Mar 2023 14:53:54	U	SMPL
CCV03 - 1	CCV03	21587	5.2116	-	7470A	PPB	01 Mar 2023 14:56:11	U	SMPL
CCB03 - 1	CCB03	516	0.0804	-	7470A	PPB	01 Mar 2023 14:58:28	U	SMPL