

LB124326


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5776E



A= 0.0000e+000

B= 1.7429e-004

C= -7.4058e-002

Rho= 0.9999874

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.006	0.006	462	0.000	462					
0.2	0.200	0.170	-0.030	1403	0.0 %	1403					-15
2.5	2.500	2.533	0.033	14959	0.0 %	14959					1
5.0	5.000	4.996	-0.004	29088	0.0 %	29088					0
7.5	7.500	7.500	-0.000	43456	0.0 %	43456					0
10.0	10.000	9.994	-0.006	57768	0.0 %	57768					0

LB124326 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Mar 2023 15:55:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	462	-	0.0000	7471B	PPB	01 Mar 2023 15:56:56	S	Std
0.2 - 1	50.2	1403	-	0.2000	7471B	PPB	01 Mar 2023 15:59:13	S	Std
2.5 - 1	52.5	14959	-	2.5000	7471B	PPB	01 Mar 2023 16:01:31	S	Std
5.0 - 1	55	29088	-	5.0000	7471B	PPB	01 Mar 2023 16:03:48	S	Std
7.5 - 1	57.5	43456	-	7.5000	7471B	PPB	01 Mar 2023 16:06:04	S	Std
10.0 - 1	510	57768	-	10.0000	7471B	PPB	01 Mar 2023 16:13:31	S	Std
ICV67 - 1	ICV67	24542	4.2034	-	7471B	PPB	01 Mar 2023 16:16:16	U	SMPL
ICB67 - 1	ICB67	198	-0.0395	-	7471B	PPB	01 Mar 2023 16:18:31	U	SMPL
CCV04 - 1	CCV04	29147	5.0060	-	7471B	PPB	01 Mar 2023 16:20:46	U	SMPL
CCB04 - 1	CCB04	180	-0.0427	-	7471B	PPB	01 Mar 2023 16:23:01	U	SMPL
CRA - 1	CRA	1404	0.1706	-	7471B	PPB	01 Mar 2023 16:25:18	U	SMPL
HighStd - 1	HighStd	58861	10.1848	-	7471B	PPB	01 Mar 2023 16:27:33	U	SMPL
ChkStd - 1	ChkStd	40590	7.0004	-	7471B	PPB	01 Mar 2023 16:29:48	U	SMPL
PB151150BL - 1	PBS	235	-0.0331	-	7471B	PPB	01 Mar 2023 16:32:06	U	SMPL
PB151150BS - 1	LCSS	22253	3.8044	-	7471B	PPB	01 Mar 2023 16:34:23	U	SMPL
O1736-01 - 1	TP-5	5408	0.8685	-	7471B	PPB	01 Mar 2023 16:36:39	U	SMPL
O1736-01DUP - 1	TP-5DUP	6018	0.9748	-	7471B	PPB	01 Mar 2023 16:38:56	U	SMPL
O1736-01MS - 1	TP-5MS	30801	5.2943	-	7471B	PPB	01 Mar 2023 16:41:13	U	SMPL
O1736-01MSD - 1	TP-5MSD	30365	5.2183	-	7471B	PPB	01 Mar 2023 16:43:30	U	SMPL
O1736-02 - 1	TP-6	4223	0.6620	-	7471B	PPB	01 Mar 2023 16:45:47	U	SMPL
CCV05 - 1	CCV05	29111	4.9997	-	7471B	PPB	01 Mar 2023 16:48:03	U	SMPL
CCB05 - 1	CCB05	164	-0.0455	-	7471B	PPB	01 Mar 2023 16:50:18	U	SMPL
O1739-01 - 1	SAMPLE-1-COMP-1	3314	0.5035	-	7471B	PPB	01 Mar 2023 16:52:34	U	SMPL
O1739-08 - 1	SAMPLE-2-COMP-2	1398	0.1696	-	7471B	PPB	01 Mar 2023 16:54:49	U	SMPL
O1739-15 - 1	SAMPLE-3-COMP-3	3156	0.4760	-	7471B	PPB	01 Mar 2023 16:57:05	U	SMPL
O1740-01 - 1	SAMPLE-4-COMP-4	1012	0.1023	-	7471B	PPB	01 Mar 2023 16:59:21	U	SMPL
O1736-01LX5 - 1		919	0.0861	-	7471B	PPB	01 Mar 2023 17:01:37	U	SMPL
O1736-01A - 1		30903	5.3120	-	7471B	PPB	01 Mar 2023 17:03:53	U	SMPL
CCV06 - 1	CCV06	29790	5.1180	-	7471B	PPB	01 Mar 2023 17:06:11	U	SMPL
CCB06 - 1	CCB06	158	-0.0465	-	7471B	PPB	01 Mar 2023 17:08:29	U	SMPL