



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB124327

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/02/2023	08:40
2	CAL2	1	Water	NA	NA	20.7	7.00	03/02/2023	08:41
3	CAL3	1	Water	NA	NA	20.8	10.02	03/02/2023	08:43
4	ICV	1	Water	NA	NA	20.6	6.98	03/02/2023	08:45
5	CCV1	1	Water	NA	NA	20.7	2.01	03/02/2023	08:47
6	O1727-04	1	Water	NA	NA	21.9	7.82	03/02/2023	08:50
7	O1743-04	1	Water	NA	NA	22.8	5.22	03/02/2023	08:55
8	O1743-04DUP	1	Water	NA	NA	22.7	5.23	03/02/2023	08:56
9	CCV2	1	Water	NA	NA	20.7	12.03	03/02/2023	08:59

VP 124328

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o1748 WorkList ID : 167800 Department : Wet-Chemistry Date : 03-02-2023 08:31:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/27/2023	Water	O1727-04	pH	Cool 4 deg C	PSEG03	E12	JB1-WATER	02/27/2023	9040C
03/01/2023	Water	O1743-04	pH	Cool 4 deg C	PSEG03	E21	34142	03/01/2023	9040C

Date/Time 03/02/23 08:35
Raw Sample Received by: JPL/WCL
Raw Sample Relinquished by: RPL C.W.L

Date/Time 03/02/23 11:00
Raw Sample Received by: RPL C.W.L
Raw Sample Relinquished by: JPL/WCL