

Analysis Method: 9034

ANALYST: Rubina

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB124333

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2941
IODINE SOLUTION .025N 1L	W2924
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB151142BL		1	50	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	03/02/2023	11:30
2	O1727-04		1	50	50	2.00	0.00	1.86	1.86	0.14	0.06	0.48	03/02/2023	11:33
3	O1727-04DUP		1	50	50	2.00	0.00	1.86	1.86	0.14	0.06	0.48	03/02/2023	11:36
4	O1748-03		1	50	50	2.00	0.00	1.88	1.88	0.12	0.04	0.32	03/02/2023	11:39
5	O1748-04		1	50	50	2.00	0.00	1.82	1.82	0.18	0.10	0.80	03/02/2023	11:42
6	O1766-02		1	50	50	2.00	0.00	1.84	1.84	0.16	0.08	0.64	03/02/2023	11:45

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume