

LAB124335


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

57673



A= 0.0000e+000

B= 1.7356e-004

C= -8.1497e-003

Rho= 0.9985323

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ (s)
0.0	0.000	-0.002	-0.002	35	0.000	35					
0.2	0.200	0.211	0.011	1261	0.0 %	1261					5
2.5	2.500	2.736	0.236	15812	0.0 %	15812					9
5.0	5.000	4.600	-0.400	26551	0.0 %	26551					-8
7.5	7.500	7.654	0.154	44148	0.0 %	44148					2
10.0	10.000	10.001	0.001	57673	0.0 %	57673					0

LB124335 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Mar 2023 12:42:45

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.0000	7471B	PPB	02 Mar 2023 13:29:57	S	Std
0.2 - 1	52	1261	-	0.2000	7471B	PPB	02 Mar 2023 13:32:15	S	Std
2.5 - 1	52.5	15812	-	2.5000	7471B	PPB	02 Mar 2023 13:34:33	S	Std
5.0 - 1	55	26551	-	5.0000	7471B	PPB	02 Mar 2023 13:39:23	S	Std
7.5 - 1	57.5	44148	-	7.5000	7471B	PPB	02 Mar 2023 13:41:40	S	Std
10.0 - 1	510	57673	-	10.0000	7471B	PPB	02 Mar 2023 13:48:28	S	Std
ICV68 - 1	ICV68	23149	4.0095	-	7471B	PPB	02 Mar 2023 13:52:43	U	SMPL
ICB68 - 1	ICB68	-64	-0.0193	-	7471B	PPB	02 Mar 2023 13:54:58	U	SMPL
CCV07 - 1	CCV07	30554	5.2947	-	7471B	PPB	02 Mar 2023 13:57:14	U	SMPL
CCB07 - 1	CCB07	-77	-0.0215	-	7471B	PPB	02 Mar 2023 13:59:29	U	SMPL
CRA - 1	CRA	1250	0.2088	-	7471B	PPB	02 Mar 2023 14:01:46	U	SMPL
HighStd - 1	HighStd	59173	10.2617	-	7471B	PPB	02 Mar 2023 14:04:01	U	SMPL
ChkStd - 1	ChkStd	40344	6.9938	-	7471B	PPB	02 Mar 2023 14:06:17	U	SMPL
PB151159BL - 1	PB151159BL	-122	-0.0293	-	7471B	PPB	02 Mar 2023 14:08:35	U	SMPL
PB151159BS - 1	PB151159BS	20421	3.5360	-	7471B	PPB	02 Mar 2023 14:10:52	U	SMPL
O1742-01 - 1	RBR251669	332	0.0495	-	7471B	PPB	02 Mar 2023 14:13:05	U	SMPL
O1744-01 - 1	72-11939	3067	0.5241	-	7471B	PPB	02 Mar 2023 14:15:26	U	SMPL
O1744-01DUP - 1	72-11939DUP	3253	0.5564	-	7471B	PPB	02 Mar 2023 14:17:43	U	SMPL
O1744-01MS - 1	72-11939MS	25500	4.4175	-	7471B	PPB	02 Mar 2023 14:20:00	U	SMPL
O1744-01MSD - 1	72-11939MSD	23208	4.0197	-	7471B	PPB	02 Mar 2023 14:22:16	U	SMPL
CCV08 - 1	CCV08	28740	4.9798	-	7471B	PPB	02 Mar 2023 14:24:32	U	SMPL
CCB08 - 1	CCB08	-84	-0.0227	-	7471B	PPB	02 Mar 2023 14:26:49	U	SMPL
O1760-01 - 1	RBR200058	1372	0.2300	-	7471B	PPB	02 Mar 2023 14:29:05	U	SMPL
O1760-03 - 1	72-11998	4938	0.8489	-	7471B	PPB	02 Mar 2023 14:31:21	U	SMPL
O1763-01 - 1	CM-A-CRUSHED-MASONRY-PILE-1	3478	0.5955	-	7471B	PPB	02 Mar 2023 14:33:37	U	SMPL
O1764-01 - 1	CS-A-CRUSHED-STONE-PILE-1	7920	1.3664	-	7471B	PPB	02 Mar 2023 14:35:53	U	SMPL
PB151160BL - 1	PB151160BL	5	-0.0073	-	7471B	PPB	02 Mar 2023 14:38:10	U	SMPL
PB151160BS - 1	PB151160BS	21235	3.6773	-	7471B	PPB	02 Mar 2023 14:43:27	U	SMPL
O1749-01 - 1	CF-3	3338	0.5712	-	7471B	PPB	02 Mar 2023 14:45:43	U	SMPL
O1749-01DUP - 1	CF-3DUP	2899	0.4950	-	7471B	PPB	02 Mar 2023 14:48:01	U	SMPL
O1749-01MS - 1	CF-3MS	25013	4.3330	-	7471B	PPB	02 Mar 2023 14:50:18	U	SMPL
O1749-01MSD - 1	CF-3MSD	25722	4.4561	-	7471B	PPB	02 Mar 2023 14:52:34	U	SMPL
CCV09 - 1	CCV09	28345	4.9113	-	7471B	PPB	02 Mar 2023 14:54:50	U	SMPL
CCB09 - 1	CCB09	-90	-0.0238	-	7471B	PPB	02 Mar 2023 14:57:06	U	SMPL
O1749-02 - 1	CF-4	3603	0.6172	-	7471B	PPB	02 Mar 2023 14:59:23	U	SMPL
O1749-03 - 1	CF-5	3752	0.6430	-	7471B	PPB	02 Mar 2023 15:01:39	U	SMPL
O1749-04 - 1	CF-6	3043	0.5200	-	7471B	PPB	02 Mar 2023 15:03:55	U	SMPL
O1749-05 - 1	CF-7	2816	0.4806	-	7471B	PPB	02 Mar 2023 15:06:11	U	SMPL
O1749-06 - 1	CF-8	3117	0.5328	-	7471B	PPB	02 Mar 2023 15:08:26	U	SMPL
O1749-07 - 1	CF-9	3210	0.5490	-	7471B	PPB	02 Mar 2023 15:10:43	U	SMPL
O1749-08 - 1	CF-10	2666	0.4546	-	7471B	PPB	02 Mar 2023 15:12:59	U	SMPL
O1749-09 - 1	CF-11	2204	0.3744	-	7471B	PPB	02 Mar 2023 15:15:17	U	SMPL
O1749-10 - 1	CF-12	3617	0.6196	-	7471B	PPB	02 Mar 2023 15:17:34	U	SMPL
O1749-11 - 1	CF-13	3019	0.5158	-	7471B	PPB	02 Mar 2023 15:19:50	U	SMPL
CCV10 - 1	CCV10	28865	5.0015	-	7471B	PPB	02 Mar 2023 15:22:06	U	SMPL
CCB10 - 1	CCB10	-79	-0.0219	-	7471B	PPB	02 Mar 2023 15:24:21	U	SMPL
O1749-12 - 1	CF-14	2998	0.5122	-	7471B	PPB	02 Mar 2023 15:26:37	U	SMPL
O1749-13 - 1	CF-15	4007	0.6873	-	7471B	PPB	02 Mar 2023 15:28:53	U	SMPL
O1749-14 - 1	CF-16	2961	0.5057	-	7471B	PPB	02 Mar 2023 15:31:09	U	SMPL
O1749-15 - 1	CF-17	3538	0.6059	-	7471B	PPB	02 Mar 2023 15:33:25	U	SMPL
O1749-16 - 1	CF-18	2976	0.5084	-	7471B	PPB	02 Mar 2023 15:35:41	U	SMPL
O1744-01LX5 - 1		560	0.0890	-	7471B	PPB	02 Mar 2023 15:37:57	U	SMPL
O1744-01A - 1		25959	4.4972	-	7471B	PPB	02 Mar 2023 15:40:14	U	SMPL
O1749-01LX5 - 1		541	0.0857	-	7471B	PPB	02 Mar 2023 15:42:31	U	SMPL
O1749-01A - 1		26381	4.5704	-	7471B	PPB	02 Mar 2023 15:44:49	U	SMPL
CCV11 - 1	CCV11	29295	5.0762	-	7471B	PPB	02 Mar 2023 15:47:05	U	SMPL
CCB11 - 1	CCB11	-66	-0.0196	-	7471B	PPB	02 Mar 2023 15:49:22	U	SMPL