



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124350

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/02/2023	16:05
2	CAL2	1	Water	NA	NA	20.7	7.00	03/02/2023	16:06
3	CAL3	1	Water	NA	NA	20.7	10.02	03/02/2023	16:08
4	ICV	1	Water	NA	NA	20.8	7.02	03/02/2023	16:10
5	CCV1	1	Water	NA	NA	20.8	2.01	03/02/2023	16:15
6	O1748-02	1	Solid	20.02	20	24.2	6.26	03/02/2023	16:20
7	O1766-01	1	Solid	20.02	20	22.8	6.37	03/02/2023	16:25
8	O1767-04	1	Solid	20.04	20	22.9	8.43	03/02/2023	16:30
9	O1767-08	1	Solid	20.02	20	23.6	9.33	03/02/2023	16:35
10	O1767-12	1	Solid	20.03	20	22.7	8.95	03/02/2023	16:38
11	O1767-16	1	Solid	20.02	20	23.5	9.06	03/02/2023	16:40
12	O1770-02	1	Solid	20.01	20	22.6	7.11	03/02/2023	16:45
13	O1770-02DUP	1	Solid	20.02	20	22.8	7.12	03/02/2023	16:46
14	CCV2	1	Water	NA	NA	20.7	12.03	03/02/2023	16:49

WORKLIST(Hardcopy Internal Chain)

VB 124350

WorkList Name : CORROSIONITY O1770 WorkList ID : 167833 Department : Wet-Chemistry Date : 03-02-2023 11:30:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage Location	Customer Sample		
03/01/2023	Solid	O1748-02	Corrosivity	Cool 4 deg C	PSEG03	E21	345	03/01/2023	9045D
03/01/2023	Solid	O1766-01	Corrosivity	Cool 4 deg C	PSEG03	E21	D2934	03/01/2023	9045D
03/01/2023	Solid	O1767-04	Corrosivity	Cool 4 deg C	RUTW01	E22	CE-45-030123	03/01/2023	9045D
03/01/2023	Solid	O1767-08	Corrosivity	Cool 4 deg C	RUTW01	E22	CE-54-030123	03/01/2023	9045D
03/01/2023	Solid	O1767-12	Corrosivity	Cool 4 deg C	RUTW01	E22	CE-63-030123	03/01/2023	9045D
03/01/2023	Solid	O1767-16	Corrosivity	Cool 4 deg C	RUTW01	E22	CE-33-030123	03/01/2023	9045D
03/02/2023	Solid	O1770-02	Corrosivity	Cool 4 deg C	PSEG03	E22	VNJ-223	03/02/2023	9045D

Date/Time 03/02/23 16:00
 Raw Sample Received by: JP(WC)
 Raw Sample Relinquished by: CP SM

Date/Time 03/02/23 17:30
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP(WC)