



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124352

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	03/02/2023	08:40
2	O1727-04		1	12	100.00	>212.0	>212.0	03/02/2023	09:10
3	O1727-04DUP		1	14	100.00	>212.0	>212.0	03/02/2023	09:40
4	O1743-04		1	14	100.00	>212.0	>212.0	03/02/2023	10:10
5	O1748-03		1	12	100.00	>212.0	>212.0	03/02/2023	10:40
6	O1748-04		1	13	100.00	>212.0	>212.0	03/02/2023	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-2-27

WorkList ID : 167692

Department : Wet-Chemistry

Date : 02-27-2023 08:41:09

LB124352

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/02/2023	Water	O1727-04	Flash Point	Cool 4 deg C	PSEG03	E12	JB1-WATER	02/27/2023	1010B
03/08/2023	Water	O1743-04	Flash Point	Cool 4 deg C	PSEG03	E21	34142	03/01/2023	1010B
03/08/2023	Water	O1748-03	Flash Point	Cool 4 deg C	PSEG03	E21	348	03/01/2023	1010B
03/08/2023	Water	O1748-04	Flash Point	Cool 4 deg C	PSEG03	E21	339	03/01/2023	1010B

Date/Time 03/02/2023 08:30

Raw Sample Received by: RM J. C. (gm)

Raw Sample Relinquished by: RM J. C. (gm)

Date/Time 03/02/2023 11:05

Raw Sample Received by: J. C. (gm)

Raw Sample Relinquished by: RM (gm)