



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124353

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	03/02/2023	12:00
2	O1748-01		1	14	100.00	>212.0	>212.0	03/02/2023	12:30
3	O1748-01DUP		1	14	100.00	>212.0	>212.0	03/02/2023	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-3-2

WorkList ID : 167802

Department : Wet-Chemistry

Date : 03-02-2023 08:41:12

LB124353

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/08/2023	Solid	O1748-01	Flash Point	Cool 4 deg C	PSEG03	E21	347	03/01/2023	1010B

Date/Time 03/02/2023 11:50
 Raw Sample Received by: RMJ (wac)
 Raw Sample Relinquished by: Del (swr)

Date/Time 03/02/2023 13:46
 Raw Sample Received by: J.C (swr)
 Raw Sample Relinquished by: RMJ (wac)