



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124363

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	03/03/2023	12:55
2	O1758-08		1	14	100.00	>212.0	>212.0	03/03/2023	13:25
3	O1758-08DUP		1	14	100.00	>212.0	>212.0	03/03/2023	13:55
4	O1758-09		1	12	100.00	>212.0	>212.0	03/03/2023	14:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66124363

WorkList Name : fp 3-3

WorkList ID : 167846

Department : Wet-Chemistry

Date : 03-03-2023 08:26:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/08/2023	Solid	O1758-08	Flash Point	Cool 4 deg C	PSEG03	E21	FMC-1221	03/01/2023	1010B
03/08/2023	Solid	O1758-09	Flash Point	Cool 4 deg C	PSEG03	E21	FMC-1222	03/01/2023	1010B

Date/Time 03/03/2023 12:45
 Raw Sample Received by: R. G. G. G.
 Raw Sample Relinquished by: R. G. G. G.

Date/Time 03/03/2023 14:40
 Raw Sample Received by: CP Jm
 Raw Sample Relinquished by: R. G. G. G.