



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124364

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	03/03/2023	09:00
2	O1758-03		1	12	100.00	>212.0	>212.0	03/03/2023	09:30
3	O1758-03DUP		1	12	100.00	>212.0	>212.0	03/03/2023	10:00
4	O1758-04		1	14	100.00	>212.0	>212.0	03/03/2023	10:30
5	O1759-01		1	14	100.00	>212.0	>212.0	03/03/2023	11:00
6	O1759-03		1	12	100.00	>212.0	>212.0	03/03/2023	11:30
7	O1766-02		1	12	100.00	>212.0	>212.0	03/03/2023	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66124364

WorkList Name : f p s 03-03 WorkList ID : 167847 Department : Wet-Chemistry Date : 03-03-2023 08:27:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/08/2023	Water	O1758-03 B	Flash Point	Cool 4 deg C	PSEG03	E21	FMC-0088	03/01/2023	1010B
03/08/2023	Water	O1758-04 B	Flash Point	Cool 4 deg C	PSEG03	E21	FMC-0089	03/01/2023	1010B
03/03/2023	Water	O1759-01 M	Flash Point	Cool 4 deg C	DVIR01	E21	DEP-1	03/01/2023	1010B
03/03/2023	Water	O1759-03 M	Flash Point	Cool 4 deg C	DVIR01	E21	DEP-2	03/01/2023	1010B
03/08/2023	Water	O1766-02 C	Flash Point	Cool 4 deg C	PSEG03	E21	D2932	03/01/2023	1010B

Date/Time 03/03/2023 08:45
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: CP M

Date/Time 03/03/2023 12:46
 Raw Sample Received by: CP M
 Raw Sample Relinquished by: RM CWC