

LB124368


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51009



A= 0.0000e+000

B= 1.9660e-004 *slope*C= -3.0722e-002 *intercept*

Rho= 0.9999472

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>4/0/0</i>
0.00	0.000	0.004	0.004	175	0.000	175					
0.05	0.050					185	0				
0.20	0.200	0.225	0.025	1303	0.0 %	1303					
2.50	2.500	2.494	-0.006	12842	0.0 %	12842					
5.00	5.000	4.928	-0.072	25221	0.0 %	25221					
7.50	7.500	7.552	0.052	38569	0.0 %	38569					
10.0	10.000	9.997	-0.003	51009	0.0 %	51009					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	175	-		0 SFAM01.1	PPB	3/5/2023 10:35	Std
	0.2 S01	1303	-		0.2 SFAM01.1	PPB	3/5/2023 10:37	Std
	2.5 S02	12842	-		2.5 SFAM01.1	PPB	3/5/2023 10:39	Std
	5 S03	25221	-		5 SFAM01.1	PPB	3/5/2023 10:42	Std
	7.5 S04	38569	-		7.5 SFAM01.1	PPB	3/5/2023 10:44	Std
	10 S05	51009	-		10 SFAM01.1	PPB	3/5/2023 10:46	Std
ICV069	ICV069	20753	4.0492	-	SFAM01.1	PPB	3/5/2023 10:52	SMPL
ICB069	ICB069	-66	-0.0437	-	SFAM01.1	PPB	3/5/2023 10:54	SMPL
CCV012	CCV012	25759	5.0334	-	SFAM01.1	PPB	3/5/2023 10:56	SMPL
CCB012	CCB012	-44	-0.0394	-	SFAM01.1	PPB	3/5/2023 10:58	SMPL
PB151180BL	PBW180	51	-0.0207	-	SFAM01.1	PPB	3/5/2023 11:01	SMPL
O1629-01	YBQJ6	50	-0.0209	-	SFAM01.1	PPB	3/5/2023 11:03	SMPL
O1629-02	YBQK2	4197	0.7944	-	SFAM01.1	PPB	3/5/2023 11:05	SMPL
O1629-03	YBQK8	474836	HIGH	-	SFAM01.1	PPB	3/5/2023 11:07	
O1629-04	YBQL0	358618	HIGH	-	SFAM01.1	PPB	3/5/2023 11:09	
O1629-05	YBQL0D	322701	HIGH	-	SFAM01.1	PPB	3/5/2023 11:11	
O1629-06	YBQL0S	366770	HIGH	-	SFAM01.1	PPB	3/5/2023 11:13	
O1629-07	YBQL6	-62	-0.0429	-	SFAM01.1	PPB	3/5/2023 11:15	SMPL
O1629-08	YBQP0	328155	HIGH	-	SFAM01.1	PPB	3/5/2023 11:17	
PB151181BL	PBW181	-321	-0.0938	-	SFAM01.1	PPB	3/5/2023 11:19	SMPL
O1630-01	YBQK3	13864	2.6949	-	SFAM01.1	PPB	3/5/2023 11:21	SMPL
O1630-02	YBQK9	449544	HIGH	-	SFAM01.1	PPB	3/5/2023 11:24	
O1630-03	YBQL1	337773	HIGH	-	SFAM01.1	PPB	3/5/2023 11:25	
O1630-04	YBQL1D	393393	77.3089	-	SFAM01.1	PPB	3/5/2023 11:27	SMPL
O1630-05	YBQL1S	350220	HIGH	-	SFAM01.1	PPB	3/5/2023 11:30	
O1630-06	YBQL7	7879	1.5183	-	SFAM01.1	PPB	3/5/2023 11:32	SMPL
O1630-07	YBQP1	303752	HIGH	-	SFAM01.1	PPB	3/5/2023 11:34	
PB151182BL	PBW182	10	-0.0288	-	SFAM01.1	PPB	3/5/2023 11:39	SMPL
O1687-01	YBQJ5	25	-0.0258	-	SFAM01.1	PPB	3/5/2023 11:41	SMPL
O1687-02	YBQK0	1233	0.2117	-	SFAM01.1	PPB	3/5/2023 11:44	SMPL
CCV013	CCV013	26006	5.082	-	SFAM01.1	PPB	3/5/2023 11:46	SMPL
CCB013	CCB013	-54	-0.0413	-	SFAM01.1	PPB	3/5/2023 11:48	SMPL
O1687-03	YBQK5	966230	HIGH	-	SFAM01.1	PPB	3/5/2023 11:50	
O1687-04	YBQL5	197833	38.8625	-	SFAM01.1	PPB	3/5/2023 11:52	SMPL
O1687-05	YBQL5D	210283	41.3102	-	SFAM01.1	PPB	3/5/2023 11:55	SMPL
O1687-06	YBQL5S	230758	45.3355	-	SFAM01.1	PPB	3/5/2023 11:57	SMPL
O1687-07	YBQN3	584280	HIGH	-	SFAM01.1	PPB	3/5/2023 11:59	
O1687-08	YBQN7	669	0.1008	-	SFAM01.1	PPB	3/5/2023 12:01	SMPL
O1687-09	YBQP5	1203	0.2058	-	SFAM01.1	PPB	3/5/2023 12:04	SMPL
PB151183BL	PBW183	-22	-0.035	-	SFAM01.1	PPB	3/5/2023 12:06	SMPL
O1688-01	YBQJ7	58	-0.0193	-	SFAM01.1	PPB	3/5/2023 12:08	SMPL
O1688-02	YBQJ8	86	-0.0138	-	SFAM01.1	PPB	3/5/2023 12:10	SMPL
O1688-03	YBQK7	1051284	HIGH	-	SFAM01.1	PPB	3/5/2023 12:13	
O1688-04	YBQK7D	1029755	HIGH	-	SFAM01.1	PPB	3/5/2023 12:14	
O1688-05	YBQK7S	1037577	HIGH	-	SFAM01.1	PPB	3/5/2023 12:17	
O1688-06	YBQL3	353160	HIGH	-	SFAM01.1	PPB	3/5/2023 12:19	

LB124368
INSTRUMENT ID : CV1

O1688-07	YBQM9	35418	6.9323 -	SFAM01.1	PPB	3/5/2023 12:21	SMPL
O1688-08	YBQN5	358610	HIGH -	SFAM01.1	PPB	3/5/2023 12:24	
O1688-09	YBQN9	33297	6.5153 -	SFAM01.1	PPB	3/5/2023 12:26	SMPL
O1688-10	YBQL8	28217	5.5166 -	SFAM01.1	PPB	3/5/2023 12:28	SMPL
O1688-11	YBQL9	552	0.0778 -	SFAM01.1	PPB	3/5/2023 12:30	SMPL
O1688-12	YBQM0	12148	2.3575 -	SFAM01.1	PPB	3/5/2023 12:33	SMPL
CCV014	CCV014	25855	5.0523 -	SFAM01.1	PPB	3/5/2023 12:35	SMPL
CCB014	CCB014	-155	-0.0612 -	SFAM01.1	PPB	3/5/2023 12:37	SMPL
PB151184BL	PBW184	58	-0.0193 -	SFAM01.1	PPB	3/5/2023 12:39	SMPL
O1689-01	YBQJ9	143	-0.0026 -	SFAM01.1	PPB	3/5/2023 12:42	SMPL
O1689-02	YBQK4	953875	HIGH -	SFAM01.1	PPB	3/5/2023 12:44	
O1689-03	YBQL4	205457	40.3614 -	SFAM01.1	PPB	3/5/2023 12:46	SMPL
O1689-04	YBQL4D	180153	35.3867 -	SFAM01.1	PPB	3/5/2023 12:48	SMPL
O1689-05	YBQL4S	204902	40.2523 -	SFAM01.1	PPB	3/5/2023 12:51	SMPL
O1689-06	YBQN2	448961	HIGH -	SFAM01.1	PPB	3/5/2023 12:53	
O1689-07	YBQN6	247	0.0178 -	SFAM01.1	PPB	3/5/2023 12:58	SMPL
O1689-08	YBQP4	79	-0.0152 -	SFAM01.1	PPB	3/5/2023 13:00	SMPL
O1689-09	YBQM1	131	-0.005 -	SFAM01.1	PPB	3/5/2023 13:02	SMPL
O1689-10	YBQM2	126	-0.006 -	SFAM01.1	PPB	3/5/2023 13:05	SMPL
O1689-11	YBQM3	252	0.0188 -	SFAM01.1	PPB	3/5/2023 13:07	SMPL
PB151185BL	PBW185	39	-0.0231 -	SFAM01.1	PPB	3/5/2023 13:09	SMPL
O1690-01	YBQK6	1059618	HIGH -	SFAM01.1	PPB	3/5/2023 13:11	
O1690-02	YBQK6D	1022395	HIGH -	SFAM01.1	PPB	3/5/2023 13:13	
O1690-03	YBQK6S	1035146	HIGH -	SFAM01.1	PPB	3/5/2023 13:16	
O1690-04	YBQL2	342665	HIGH -	SFAM01.1	PPB	3/5/2023 13:18	
O1690-05	YBQM8	-35	-0.0376 -	SFAM01.1	PPB	3/5/2023 13:23	SMPL
O1690-06	YBQN4	348553	HIGH -	SFAM01.1	PPB	3/5/2023 13:25	
O1690-07	YBQN8	664	0.0998 -	SFAM01.1	PPB	3/5/2023 13:27	SMPL
CCV015	CCV015	25247	4.9327 -	SFAM01.1	PPB	3/5/2023 13:29	SMPL
CCB015	CCB015	-52	-0.0409 -	SFAM01.1	PPB	3/5/2023 13:31	SMPL
O1629-03DLX25		29632	5.7948 -	SFAM01.1	PPB	3/5/2023 13:34	SMPL
O1629-04DLX25		20163	3.9333 -	SFAM01.1	PPB	3/5/2023 13:36	SMPL
O1629-05DLX25		21496	4.1953 -	SFAM01.1	PPB	3/5/2023 13:44	SMPL
O1629-06DLX25		19797	3.8613 -	SFAM01.1	PPB	3/5/2023 13:47	SMPL
O1629-08DLX25		16453	3.2039 -	SFAM01.1	PPB	3/5/2023 13:49	SMPL
O1630-02DLX25		22689	4.4299 -	SFAM01.1	PPB	3/5/2023 13:54	SMPL
O1630-03DLX25		16949	3.3014 -	SFAM01.1	PPB	3/5/2023 13:56	SMPL
O1630-05DLX25		15951	3.1052 -	SFAM01.1	PPB	3/5/2023 14:01	SMPL
O1630-07DLX25		14835	2.8858 -	SFAM01.1	PPB	3/5/2023 14:03	SMPL
O1687-03DLX10		85426	16.7637 -	SFAM01.1	PPB	3/5/2023 14:06	SMPL
O1687-04DLX10		20701	4.039 -	SFAM01.1	PPB	3/5/2023 14:08	SMPL
O1687-05DLX10		19026	3.7097 -	SFAM01.1	PPB	3/5/2023 14:10	SMPL
O1687-07DLX10		92572	18.1686 -	SFAM01.1	PPB	3/5/2023 14:15	SMPL
CCV016	CCV016	25682	5.0183 -	SFAM01.1	PPB	3/5/2023 14:17	SMPL
CCB016	CCB016	-85	-0.0474 -	SFAM01.1	PPB	3/5/2023 14:20	SMPL
O1688-03DLX10		231239	45.43 -	SFAM01.1	PPB	3/5/2023 14:22	SMPL
O1688-04DLX10		228526	44.8967 -	SFAM01.1	PPB	3/5/2023 14:24	SMPL

LB124368
INSTRUMENT ID : CV1

O1688-05DLX10		228404	44.8727 -	SFAM01.1	PPB	3/5/2023 14:26	SMPL
O1688-08DLX25		19280	3.7597 -	SFAM01.1	PPB	3/5/2023 14:34	SMPL
O1689-02DLX10		186781	36.6898 -	SFAM01.1	PPB	3/5/2023 14:36	SMPL
O1689-03DLX10		18693	3.6443 -	SFAM01.1	PPB	3/5/2023 14:39	SMPL
O1689-04DLX10		18662	3.6382 -	SFAM01.1	PPB	3/5/2023 14:44	SMPL
O1689-05DLX10		20083	3.9175 -	SFAM01.1	PPB	3/5/2023 14:46	SMPL
O1689-06DLX25		19973	3.8959 -	SFAM01.1	PPB	3/5/2023 14:48	SMPL
O1690-01DLX10		238339	46.8259 -	SFAM01.1	PPB	3/5/2023 14:51	SMPL
O1690-02DLX10		224168	44.0399 -	SFAM01.1	PPB	3/5/2023 14:53	SMPL
O1690-03DLX10		222612	43.734 -	SFAM01.1	PPB	3/5/2023 14:55	SMPL
O1690-04DLX25		17690	3.4471 -	SFAM01.1	PPB	3/5/2023 14:58	SMPL
O1690-06DLX25		20137	3.9281 -	SFAM01.1	PPB	3/5/2023 15:00	SMPL
CCV017	CCV017	25951	5.0712 -	SFAM01.1	PPB	3/5/2023 15:02	SMPL
CCB017	CCB017	-73	-0.0451 -	SFAM01.1	PPB	3/5/2023 15:05	SMPL
O1687-07DL2X50		16378	3.1891 -	SFAM01.1	PPB	3/5/2023 15:09	SMPL
O1688-03DL2X100		21099	4.1173 -	SFAM01.1	PPB	3/5/2023 15:11	SMPL
O1688-04DL2X100		22832	4.458 -	SFAM01.1	PPB	3/5/2023 15:17	SMPL
O1688-05DL2X100		22742	4.4403 -	SFAM01.1	PPB	3/5/2023 15:19	SMPL
O1689-02DL2X100		17712	3.4514 -	SFAM01.1	PPB	3/5/2023 15:22	SMPL
O1690-01DL2X100		23178	4.526 -	SFAM01.1	PPB	3/5/2023 15:24	SMPL
O1690-02DL2X100		22637	4.4196 -	SFAM01.1	PPB	3/5/2023 15:29	SMPL
O1690-03DL2X100		21416	4.1796 -	SFAM01.1	PPB	3/5/2023 15:31	SMPL
O1630-04DLX25		19799	3.8617 -	SFAM01.1	PPB	3/5/2023 15:34	SMPL
O1687-06DLX10		22715	4.435 -	SFAM01.1	PPB	3/5/2023 15:36	SMPL
O1688-06DLX25		22629	4.4181 -	SFAM01.1	PPB	3/5/2023 15:47	SMPL
O1687-03DL2X50		19466	3.7962 -	SFAM01.1	PPB	3/5/2023 15:50	SMPL
CCV018	CCV018	25913	5.0637 -	SFAM01.1	PPB	3/5/2023 15:52	SMPL
CCB018	CCB018	-77	-0.0459 -	SFAM01.1	PPB	3/5/2023 15:54	SMPL