



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124371

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	03/06/2023	11:15
2	O1798-03		1	14	100.00	>212.0	>212.0	03/06/2023	11:45
3	O1798-03DUP		1	14	100.00	>212.0	>212.0	03/06/2023	12:15
4	O1803-03		1	12	100.00	>212.0	>212.0	03/06/2023	12:45
5	O1803-04		1	12	100.00	>212.0	>212.0	03/06/2023	13:15
6	O1803-05		1	14	100.00	>212.0	>212.0	03/06/2023	13:45
7	O1803-07		1	13	100.00	>212.0	>212.0	03/06/2023	14:15
8	O1803-08		1	14	100.00	>212.0	>212.0	03/06/2023	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6124371

WorkList Name : fp-3-6

WorkList ID : 167900

Department : Wet-Chemistry

Date : 03-06-2023 08:35:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/10/2023	Water	O1798-03	Flash Point	Cool 4 deg C	PSEG03	E31	TOTE-COMP-1	03/03/2023	1010B
03/10/2023	Water	O1803-03	Flash Point	Cool 4 deg C	PSEG03	E31	31906	03/03/2023	1010B
03/10/2023	Water	O1803-04	Flash Point	Cool 4 deg C	PSEG03	E31	31603	03/03/2023	1010B
03/10/2023	Water	O1803-05	Flash Point	Cool 4 deg C	PSEG03	E31	COMP	03/03/2023	1010B
03/10/2023	Water	O1803-07	Flash Point	Cool 4 deg C	PSEG03	E31	1710	03/03/2023	1010B
03/10/2023	Water	O1803-08	Flash Point	Cool 4 deg C	PSEG03	E31	1711	03/03/2023	1010B

Date/Time 03/06/2023 11:05
 Raw Sample Received by: RM CWO
 Raw Sample Relinquished by: JH CWO

Date/Time 03/06/2023 15:30
 Raw Sample Received by: JH CWO
 Raw Sample Relinquished by: RM CWO