

LB 124375


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

45443



A= 0.0000e+000

B= 2.1909e-004

C= -4.5048e-002

Rho= 0.9998782

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	7.1)
0.0	0.000	-0.033	-0.033	53	0.000	53					-11
0.2	0.200	0.179	-0.021	1021	0.0 %	1021					-1
2.5	2.500	2.533	0.033	11769	0.0 %	11769					0
5.0	5.000	5.015	0.015	23095	0.0 %	23095					1
7.5	7.500	7.595	0.095	34873	0.0 %	34873					1
10.0	10.000	9.911	-0.089	45443	0.0 %	45443					-1

LB124375 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Mar 2023 10:56:41

Sample ID	Extended ID	Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	53	-	0.0007470A	PPB		06 Mar 2023 11:21:09	S	Std
0.2 - 1	50.2	1021	-	0.20007470A	PPB		06 Mar 2023 11:23:27	S	Std
2.5 - 1	52.5	11769	-	2.50007470A	PPB		06 Mar 2023 11:25:44	S	Std
5.0 - 1	55	23095	-	5.00007470A	PPB		06 Mar 2023 11:28:02	S	Std
7.5 - 1	57.5	34873	-	7.50007470A	PPB		06 Mar 2023 11:32:57	S	Std
10.0 - 1	510	45443	-	10.00007470A	PPB		06 Mar 2023 11:37:49	S	Std
ICV71 - 1	ICV71	19025	4.1232	-7470A	PPB		06 Mar 2023 11:43:10	U	SMPL
ICB71 - 1	ICB71	-55	-0.0571	-7470A	PPB		06 Mar 2023 11:45:26	U	SMPL
CCV19 - 1	CCV19	24652	5.3560	-7470A	PPB		06 Mar 2023 11:47:40	U	SMPL
CCB19 - 1	CCB19	-76	-0.0617	-7470A	PPB		06 Mar 2023 11:49:56	U	SMPL
CRA - 1	CRA	1125	0.2014	-7470A	PPB		06 Mar 2023 11:54:46	U	SMPL
HighStd - 1	HighStd	46341	9.8888	-7470A	PPB		06 Mar 2023 11:57:01	U	SMPL
ChkStd - 1	ChkStd	32599	7.0971	-7470A	PPB		06 Mar 2023 11:59:16	U	SMPL
PB151210BL - 1	PBW	-119	-0.0711	-7470A	PPB		06 Mar 2023 12:01:34	U	SMPL
PB151210BS - 1	LCSW	15590	3.3706	-7470A	PPB		06 Mar 2023 12:03:51	U	SMPL
O1748-01 - 1	347	3	-0.0444	-7470A	PPB		06 Mar 2023 12:06:08	U	SMPL
O1748-02 - 1	345	46	-0.0350	-7470A	PPB		06 Mar 2023 12:08:24	U	SMPL
O1766-01 - 1	D2934	65	-0.0308	-7470A	PPB		06 Mar 2023 12:10:41	U	SMPL
O1767-04 - 1	CE-45-030123	312	0.0233	-7470A	PPB		06 Mar 2023 12:12:58	U	SMPL
O1767-08 - 1	CE-54-030123	34	-0.0376	-7470A	PPB		06 Mar 2023 12:16:14	U	SMPL
CCV20 - 1	CCV20	24947	5.4206	-7470A	PPB		06 Mar 2023 12:17:29	U	SMPL
CCB20 - 1	CCB20	-54	-0.0569	-7470A	PPB		06 Mar 2023 12:19:44	U	SMPL
O1767-12 - 1	CE-63-030123	75	-0.0286	-7470A	PPB		06 Mar 2023 12:22:00	U	SMPL
O1767-16 - 1	CE-33-030123	81	-0.0273	-7470A	PPB		06 Mar 2023 12:24:15	U	SMPL
O1770-02 - 1	VNJ-223	40	-0.0363	-7470A	PPB		06 Mar 2023 12:26:31	U	SMPL
O1770-02DUP - 1	VNJ-223DUP	170	-0.0078	-7470A	PPB		06 Mar 2023 12:28:47	U	SMPL
O1770-02MS - 1	VNJ-223MS	16303	3.6268	-7470A	PPB		06 Mar 2023 12:31:03	U	SMPL
O1770-02MSD - 1	VNJ-223MSD	15954	3.4503	-7470A	PPB		06 Mar 2023 12:33:20	U	SMPL
PB151211BL - 1	PBW	-41	-0.0540	-7470A	PPB		06 Mar 2023 12:35:37	U	SMPL
PB151211BS - 1	LCSW	15943	3.4479	-7470A	PPB		06 Mar 2023 12:37:54	U	SMPL
O1748-03 - 1	348	-20	-0.0494	-7470A	PPB		06 Mar 2023 12:40:11	U	SMPL
O1748-03DUP - 1	348DUP	15	-0.0418	-7470A	PPB		06 Mar 2023 12:42:26	U	SMPL
CCV21 - 1	CCV21	24954	5.4222	-7470A	PPB		06 Mar 2023 12:44:43	U	SMPL
CCB21 - 1	CCB21	-80	-0.0626	-7470A	PPB		06 Mar 2023 12:46:58	U	SMPL
O1748-03MS - 1	348MS	16887	3.6548	-7470A	PPB		06 Mar 2023 12:49:16	U	SMPL
O1748-03MSD - 1	348MSD	16137	3.4904	-7470A	PPB		06 Mar 2023 12:51:32	U	SMPL
O1748-04 - 1	339	28	-0.0389	-7470A	PPB		06 Mar 2023 12:53:47	U	SMPL
O1766-02 - 1	D2932	62	-0.0315	-7470A	PPB		06 Mar 2023 12:56:03	U	SMPL
PB151209BL - 1	PBW	-3	-0.0457	-7470A	PPB		06 Mar 2023 12:58:20	U	SMPL
PB151209BS - 1	LCSW	17531	3.7958	-7470A	PPB		06 Mar 2023 13:00:36	U	SMPL
O1741-02 - 1	COMP	262	0.0124	-7470A	PPB		06 Mar 2023 13:02:54	U	SMPL
O1759-01 - 1	DEP-1	349922	76.6200	-7470A	PPB		06 Mar 2023 13:05:11	U	SMPL
O1759-03 - 1	DEP-2	29837	6.4920	-7470A	PPB		06 Mar 2023 13:07:28	U	SMPL
O1759-03DUP - 1	DEP-2DUP	33196	7.2279	-7470A	PPB		06 Mar 2023 13:10:00	U	SMPL
CCV22 - 1	CCV22	22809	4.9522	-7470A	PPB		06 Mar 2023 13:12:16	U	SMPL
CCB22 - 1	CCB22	-88	-0.0643	-7470A	PPB		06 Mar 2023 13:14:32	U	SMPL
O1759-03MS - 1	DEP-2MS	52375	11.4299	-7470A	PPB		06 Mar 2023 13:16:48	U	SMPL
O1759-03MSD - 1	DEP-2MSD	50282	10.9713	-7470A	PPB		06 Mar 2023 13:19:04	U	SMPL
PB151168TB - 1	PB151168TB	-125	-0.0724	-7470A	PPB		06 Mar 2023 13:21:22	U	SMPL
PB151168TB - 1	PB151168TB	32	-0.0380	-7470A	PPB		06 Mar 2023 13:23:39	U	SMPL
O1770-02LX5 - 1		-25	-0.0505	-7470A	PPB		06 Mar 2023 13:25:55	U	SMPL
O1770-02A - 1		17224	3.7286	-7470A	PPB		06 Mar 2023 13:28:12	U	SMPL
O1748-03LX5 - 1		-80	-0.0626	-7470A	PPB		06 Mar 2023 13:30:28	U	SMPL
O1748-03A - 1		19110	4.1418	-7470A	PPB		06 Mar 2023 13:32:45	U	SMPL
O1759-03LX5 - 1		6608	1.4027	-7470A	PPB		06 Mar 2023 13:35:03	U	SMPL
O1759-03A - 1		51809	11.3059	-7470A	PPB		06 Mar 2023 13:37:19	U	SMPL
CCV23 - 1	CCV23	23179	5.0333	-7470A	PPB		06 Mar 2023 13:39:35	U	SMPL
CCB23 - 1	CCB23	-83	-0.0632	-7470A	PPB		06 Mar 2023 13:41:52	U	SMPL
O1759-01DLX10 - 1		0	0.0450	-7470A	PPB		06 Mar 2023 13:44:08	U	SMPL
O1759-01DLX10 - 1		44921	9.7968	-7470A	PPB		06 Mar 2023 13:58:04	U	SMPL
O1759-03DLX2 - 1		15475	3.3454	-7470A	PPB		06 Mar 2023 14:00:20	U	SMPL
O1759-03DLX2 - 1		17664	3.8250	-7470A	PPB		06 Mar 2023 14:02:37	U	SMPL
O1759-03DLX2 - 1		27817	6.0494	-7470A	PPB		06 Mar 2023 14:04:53	U	SMPL
O1759-03DLX2 - 1		26852	5.8380	-7470A	PPB		06 Mar 2023 14:07:09	U	SMPL
O1759-03DLX10 - 1		2963	0.6041	-7470A	PPB		06 Mar 2023 14:09:25	U	SMPL
O1759-03ADLX2 - 1		26745	5.8146	-7470A	PPB		06 Mar 2023 14:11:42	U	SMPL
CCV24 - 1	CCV24	22741	4.9373	-7470A	PPB		06 Mar 2023 14:13:59	U	SMPL
CCB24 - 1	CCB24	-66	-0.0595	-7470A	PPB		06 Mar 2023 14:16:17	U	SMPL