



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 3/7/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 03/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 03/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB124376

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1805-01	EYFB0	1	1.12	8.77	9.89	8.09	79.5	
O1805-02	EYFB1	2	1.16	8.50	9.66	8.15	82.2	
O1805-03	EYFH7	3	1.18	8.49	9.67	8.28	83.6	
O1805-04	EYFH8	4	1.19	8.41	9.6	8.37	85.4	
O1805-05	EYFH9	5	1.17	8.50	9.67	8.42	85.3	
O1805-06	EYFJ0	6	1.15	8.81	9.96	8.27	80.8	
O1805-07	EYFJ1	7	1.15	8.77	9.92	8.81	87.3	
O1805-08	EYFJ2	8	1.17	8.81	9.98	8.56	83.9	
O1805-09	EYFJ3	9	1.17	8.73	9.9	8.41	82.9	
O1805-10	EYFJ4	10	1.1	8.77	9.87	8.41	83.4	
O1805-11	EYFJ4MS	11	1.1	8.77	9.87	8.41	83.4	
O1805-12	EYFJ4MSD	12	1.1	8.77	9.87	8.41	83.4	
O1805-13	EYFJ5	13	1.13	8.46	9.59	8.13	82.7	
O1805-14	EYFJ6	14	1.12	8.86	9.98	7.61	73.3	
O1805-15	EYFJ7	15	1.18	8.43	9.61	7.77	78.2	
O1805-16	EYFJ8	16	1.19	8.45	9.64	7.93	79.8	
O1805-17	EYFJ9	17	1.18	8.79	9.97	8.86	87.4	
O1805-18	EYFK0	18	1.15	8.82	9.97	8.82	87.0	
O1805-19	EYFK1	19	1.19	8.60	9.79	8.27	82.3	
O1805-20	EYFK2	20	1.14	8.53	9.67	7.84	78.5	
O1805-21	EYFK3	21	1.16	8.65	9.81	8.21	81.5	
O1805-22	EYFK4	22	1.19	8.46	9.65	8.17	82.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$