



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 3/7/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:05
In Date: 03/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:28
Out Date: 03/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB124378

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1807-01	EYFM5	1	1.18	8.54	9.72	8.11	81.1	
O1807-02	EYFM6	2	1.19	8.57	9.76	8.28	82.7	
O1807-03	EYFM7	3	1.18	8.79	9.97	8.72	85.8	
O1807-04	EYFM8	4	1.18	8.78	9.96	8.49	83.3	
O1807-05	EYFM9	5	1.19	8.63	9.82	8.13	80.4	
O1807-06	EYFN0	6	1.19	8.57	9.76	8.2	81.8	
O1807-07	EYFN1	7	1.17	8.65	9.82	8.67	86.7	
O1807-08	EYFN2	8	1.18	8.52	9.7	8.76	89.0	
O1807-09	EYFN3	9	1.19	8.74	9.93	8.91	88.3	
O1807-10	EYFN4	10	1.15	8.79	9.94	8.93	88.5	
O1807-11	EYFN4MS	11	1.15	8.79	9.94	8.93	88.5	
O1807-12	EYFN4MSD	12	1.15	8.79	9.94	8.93	88.5	
O1807-13	EYFN5	13	1.13	8.66	9.79	8.79	88.5	
O1807-14	EYFN6	14	1.18	8.48	9.66	7.81	78.2	
O1807-15	EYFN7	15	1.13	8.46	9.59	8.13	82.7	
O1807-16	EYFN8	16	1.1	8.87	9.97	8.52	83.7	
O1807-17	EYFN9	17	1.15	8.81	9.96	8.92	88.2	
O1807-18	EYFP0	18	1.13	8.42	9.55	8.33	85.5	
O1807-19	EYFP1	19	1.16	8.81	9.97	8.35	81.6	
O1807-20	EYFP2	20	1.19	8.51	9.7	8.6	87.1	
O1807-21	EYFP3	21	1.16	8.81	9.97	8.91	88.0	
O1807-22	EYFP4	22	1.12	8.81	9.93	8.71	86.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$