

LB124382


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

63675



A= 0.0000e+000

B= 1.5655e-004

C= -8.9345e-002

Rho= 0.9996779

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/bd
0.0	0.000	-0.018	-0.018	455	0.000	455					-10
0.2	0.200	0.181	-0.019	1726	0.0 %	1726					1
2.5	2.500	2.533	0.033	16754	0.0 %	16754					-1
5.0	5.000	4.943	-0.057	32147	0.0 %	32147					2
7.5	7.500	7.681	0.181	49638	0.0 %	49638					-1
10.0	10.000	9.879	-0.121	63675	0.0 %	63675					

LB124382 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Mar 2023 15:27:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	455	-	0.0000	7471B	PPB	06 Mar 2023 15:48:08	S	Std
0.2 - 1	50.2	1726	-	0.2000	7471B	PPB	06 Mar 2023 15:50:26	S	Std
2.5 - 1	52.5	16754	-	2.5000	7471B	PPB	06 Mar 2023 15:52:43	S	Std
5.0 - 1	55	32147	-	5.0000	7471B	PPB	06 Mar 2023 15:55:00	S	Std
7.5 - 1	57.5	49638	-	7.5000	7471B	PPB	06 Mar 2023 15:57:16	S	Std
10.0 - 1	59	63675	-	10.0000	7471B	PPB	06 Mar 2023 16:02:11	S	Std
ICV72 - 1	ICV72	27896	4.2778	-	7471B	PPB	06 Mar 2023 16:04:56	U	SMPL
ICB72 - 1	ICB72	440	-0.0205	-	7471B	PPB	06 Mar 2023 16:07:11	U	SMPL
CCV25 - 1	CCV25	32283	4.9646	-	7471B	PPB	06 Mar 2023 16:09:27	U	SMPL
CCB25 - 1	CCB25	435	-0.0212	-	7471B	PPB	06 Mar 2023 16:11:42	U	SMPL
CRA - 1	CRA	1737	0.1826	-	7471B	PPB	06 Mar 2023 16:13:58	U	SMPL
HighStd - 1	HighStd	65283	10.1307	-	7471B	PPB	06 Mar 2023 16:16:13	U	SMPL
ChkStd - 1	ChkStd	45904	7.0969	-	7471B	PPB	06 Mar 2023 16:18:29	U	SMPL
PB151251BL - 1	PBS	335	-0.0369	-	7471B	PPB	06 Mar 2023 16:20:47	U	SMPL
PB151251BS - 1	LCSS	23363	3.5681	-	7471B	PPB	06 Mar 2023 16:23:04	U	SMPL
O1767-03 - 1	CE-45-030123	48915	7.5683	-	7471B	PPB	06 Mar 2023 16:25:21	U	SMPL
O1767-07 - 1	CE-54-030123	2900	0.3647	-	7471B	PPB	06 Mar 2023 16:27:39	U	SMPL
O1767-11 - 1	CE-63-030123	113253	17.6404	-	7471B	PPB	06 Mar 2023 16:29:58	U	SMPL
O1767-15 - 1	CE-33-030123	24312	3.7167	-	7471B	PPB	06 Mar 2023 16:32:15	U	SMPL
O1770-01 - 1	VNJ-223	3940	0.5275	-	7471B	PPB	06 Mar 2023 16:34:33	U	SMPL
CCV26 - 1	CCV26	32287	4.9652	-	7471B	PPB	06 Mar 2023 16:36:50	U	SMPL
CCB26 - 1	CCB26	433	-0.0216	-	7471B	PPB	06 Mar 2023 16:39:05	U	SMPL
O1776-01 - 1	TR-6-030223	17617	2.6686	-	7471B	PPB	06 Mar 2023 16:41:21	U	SMPL
O1787-01 - 1	COMP-1	15362	2.3156	-	7471B	PPB	06 Mar 2023 16:48:48	U	SMPL
O1787-01DUP - 1	COMP-1DUP	16862	2.5504	-	7471B	PPB	06 Mar 2023 16:51:03	U	SMPL
O1787-01MS - 1	COMP-1MS	41515	6.4098	-	7471B	PPB	06 Mar 2023 16:56:05	U	SMPL
O1787-01MSD - 1	COMP-1MSD	41380	6.3887	-	7471B	PPB	06 Mar 2023 16:58:21	U	SMPL
O1787-02 - 1	COMP-2	2021	0.2270	-	7471B	PPB	06 Mar 2023 17:00:39	U	SMPL
O1787-03 - 1	COMP-3	19623	2.9826	-	7471B	PPB	06 Mar 2023 17:02:57	U	SMPL
O1814-02 - 1	VNJ-206	2699	0.3332	-	7471B	PPB	06 Mar 2023 17:05:14	U	SMPL
O1815-01 - 1	OK-01-03032023	2429	0.2909	-	7471B	PPB	06 Mar 2023 17:07:33	U	SMPL
O1817-01 - 1	PL-04-03032023	2143	0.2461	-	7471B	PPB	06 Mar 2023 17:09:49	U	SMPL
CCV27 - 1	CCV27	32115	4.9383	-	7471B	PPB	06 Mar 2023 17:12:05	U	SMPL
CCB27 - 1	CCB27	388	-0.0286	-	7471B	PPB	06 Mar 2023 17:14:20	U	SMPL
O1820-01 - 1	HR-3-03032023	9954	1.4690	-	7471B	PPB	06 Mar 2023 17:16:38	U	SMPL
O1820-03 - 1	HR-4-03032023	6484	0.9257	-	7471B	PPB	06 Mar 2023 17:18:53	U	SMPL
O1787-01LX5 - 1		3663	0.4841	-	7471B	PPB	06 Mar 2023 17:21:09	U	SMPL
O1787-01A - 1		41461	6.4014	-	7471B	PPB	06 Mar 2023 17:23:24	U	SMPL
O1767-11DLX5 - 1		24502	3.7464	-	7471B	PPB	06 Mar 2023 17:25:40	U	SMPL
CCV28 - 1	CCV28	32287	4.9652	-	7471B	PPB	06 Mar 2023 17:28:08	U	SMPL
CCB28 - 1	CCB28	450	-0.0189	-	7471B	PPB	06 Mar 2023 17:30:31	U	SMPL