



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB124384

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/06/2023	16:10
2	CAL2	1	Water	NA	NA	20.7	7.00	03/06/2023	16:11
3	CAL3	1	Water	NA	NA	20.8	10.03	03/06/2023	16:12
4	ICV	1	Water	NA	NA	20.7	6.99	03/06/2023	16:15
5	CCV1	1	Water	NA	NA	20.7	2.01	03/06/2023	16:20
6	O1798-01	1	Solid	20.02	20	23.4	8.55	03/06/2023	16:25
7	O1821-02	1	Solid	20.03	20	22.6	7.60	03/06/2023	16:28
8	O1825-02	1	Solid	20.01	20	23.1	7.94	03/06/2023	16:30
9	O1825-04	1	Solid	20.02	20	22.7	7.68	03/06/2023	16:33
10	O1825-06	1	Solid	20.03	20	22.7	7.41	03/06/2023	16:35
11	O1825-08	1	Solid	20.02	20	22.9	7.52	03/06/2023	16:40
12	O1825-08DUP	1	Solid	20.03	20	22.8	7.53	03/06/2023	16:41
13	CCV2	1	Water	NA	NA	20.7	12.03	03/06/2023	16:42

WORKLIST(Hardcopy Internal Chain)

18124384

WorkList Name : CORROSIONITY O1825

WorkList ID : 167929

Department : Wet-Chemistry

Date : 03-06-2023 15:39:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/03/2023	Solid	O1798-01	Corrosivity	Cool 4 deg C	PSEG03	E31	TOTE-2	03/03/2023	9045D
03/03/2023	Solid	O1821-02	Corrosivity	Cool 4 deg C	PORT06	E31	WC-IDW-S-1	03/03/2023	9045D
03/06/2023	Solid	O1825-02	Corrosivity	Cool 4 deg C	PSEG03	E31	HH-TP1-0306	03/06/2023	9045D
03/06/2023	Solid	O1825-04	Corrosivity	Cool 4 deg C	PSEG03	E31	HH-TP2-0306	03/06/2023	9045D
03/06/2023	Solid	O1825-06	Corrosivity	Cool 4 deg C	PSEG03	E31	HH-TP3-0306	03/06/2023	9045D
03/06/2023	Solid	O1825-08	Corrosivity	Cool 4 deg C	PSEG03	E31	HH-TP4-0306	03/06/2023	9045D

18145
 Date/Time 03/06/23 15:30 SP
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

17:30
 Date/Time 03/06/23
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]