



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB124389

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	03/07/2023	12:00
2	O1816-01		1	14	100.00	>212.0	>212.0	03/07/2023	12:30
3	O1816-02		1	12	100.00	>212.0	>212.0	03/07/2023	13:00
4	O1816-03		1	12	100.00	>212.0	>212.0	03/07/2023	13:30
5	O1816-05		1	14	100.00	>212.0	>212.0	03/07/2023	14:00
6	O1816-06		1	14	100.00	>212.0	>212.0	03/07/2023	14:30
7	O1816-07		1	12	100.00	>212.0	>212.0	03/07/2023	15:00
8	O1816-07DUP		1	12	100.00	>212.0	>212.0	03/07/2023	15:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66124389

WorkList Name : fp3-7

WorkList ID : 167937

Department : Wet-Chemistry

Date : 03-07-2023 08:24:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/10/2023	Water	O1816-01	Flash Point	Cool 4 deg C	PSEG03	E31	2321	03/03/2023	1010B
03/10/2023	Water	O1816-02	Flash Point	Cool 4 deg C	PSEG03	E31	91	03/03/2023	1010B
03/10/2023	Water	O1816-03	Flash Point	Cool 4 deg C	PSEG03	E31	COMP	03/03/2023	1010B
03/10/2023	Water	O1816-05	Flash Point	Cool 4 deg C	PSEG03	E31	31957	03/03/2023	1010B
03/10/2023	Water	O1816-06	Flash Point	Cool 4 deg C	PSEG03	E31	2325	03/03/2023	1010B
03/10/2023	Water	O1816-07	Flash Point	Cool 4 deg C	PSEG03	E31	662	03/03/2023	1010B

Date/Time 03/07/2023 11:40
 Raw Sample Received by: RM
 Raw Sample Relinquished by: RM

Date/Time 03/07/2023 16:10
 Raw Sample Received by: RM
 Raw Sample Relinquished by: RM