

LA124394


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

4391E



A= 0.0000e+000

B= 2.2680e-004

C= -2.2799e-002

Rho= 0.9998155

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.0	0.000	-0.005	-0.005	77	0.000	77					
0.2	0.200	0.217	0.017	1059	0.0 %	1059					
2.5	2.500	2.500	0.000	11124	0.0 %	11124					
5.0	5.000	4.914	-0.086	21769	0.0 %	21769					
7.5	7.500	7.636	0.136	33770	0.0 %	33770					
10.0	10.000	9.937	-0.063	43916	0.0 %	43916					

1000
1200
1200
1200

LB124394 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Mar 2023 13:29:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	77	-	0.0000	7470A	PPB	07 Mar 2023 13:46:27	S	Std
0.2 - 1	50	1059	-	0.2000	7470A	PPB	07 Mar 2023 13:48:45	S	Std
2.5 - 1	52	11124	-	2.5000	7470A	PPB	07 Mar 2023 13:51:02	S	Std
5.0 - 1	55	21769	-	5.0000	7470A	PPB	07 Mar 2023 13:53:19	S	Std
7.5 - 1	52.5	33770	-	7.5000	7470A	PPB	07 Mar 2023 13:55:36	S	Std
10.0 - 1	51V	43916	-	10.0000	7470A	PPB	07 Mar 2023 14:00:25	S	Std
ICV73 - 1	ICV73	18317	4.1315	-7470A	PPB	07 Mar 2023 14:03:08	U	SMPL	
ICB73 - 1	ICB73	-8	-0.0246	-7470A	PPB	07 Mar 2023 14:05:23	U	SMPL	
CCV29 - 1	CCV29	21470	4.8466	-7470A	PPB	07 Mar 2023 14:07:38	U	SMPL	
CCB29 - 1	CCB29	7	-0.0212	-7470A	PPB	07 Mar 2023 14:09:53	U	SMPL	
CRA - 1	CRA	1079	0.2219	-7470A	PPB	07 Mar 2023 14:12:08	U	SMPL	
HighStd - 1	HighStd	43603	9.8663	-7470A	PPB	07 Mar 2023 14:14:24	U	SMPL	
ChkStd - 1	ChkStd	30916	6.9889	-7470A	PPB	07 Mar 2023 14:16:40	U	SMPL	
PB151273BL - 1	PBW	-73	-0.0394	-7470A	PPB	07 Mar 2023 14:18:57	U	SMPL	
PB151273BS - 1	LCSW	15409	3.4719	-7470A	PPB	07 Mar 2023 14:21:16	U	SMPL	
O1798-01 - 1	TOTE-2	9	-0.0208	-7470A	PPB	07 Mar 2023 14:23:34	U	SMPL	
O1821-02 - 1	WC-IDW-S-1	38	-0.0142	-7470A	PPB	07 Mar 2023 14:25:52	U	SMPL	
O1825-02 - 1	HH-TP1-0306	61	-0.0090	-7470A	PPB	07 Mar 2023 14:28:09	U	SMPL	
O1825-04 - 1	HH-TP2-0306	63	-0.0085	-7470A	PPB	07 Mar 2023 14:30:27	U	SMPL	
O1825-06 - 1	HH-TP3-0306	64	-0.0083	-7470A	PPB	07 Mar 2023 14:32:42	U	SMPL	
CCV30 - 1	CCV30	22366	5.0498	-7470A	PPB	07 Mar 2023 14:34:58	U	SMPL	
CCB30 - 1	CCB30	-26	-0.0287	-7470A	PPB	07 Mar 2023 14:37:13	U	SMPL	
O1825-08 - 1	HH-TP4-0306	77	-0.0053	-7470A	PPB	07 Mar 2023 14:39:29	U	SMPL	
O1825-08DUP - 1	HH-TP4-0306DUP	47	-0.0121	-7470A	PPB	07 Mar 2023 14:41:44	U	SMPL	
O1825-08MS - 1	HH-TP4-0306MS	16733	3.7722	-7470A	PPB	07 Mar 2023 14:44:00	U	SMPL	
O1825-08MSD - 1	HH-TP4-0306MSD	16548	3.7303	-7470A	PPB	07 Mar 2023 14:46:16	U	SMPL	
PB151274BL - 1	PBW	-4	-0.0237	-7470A	PPB	07 Mar 2023 14:48:32	U	SMPL	
PB151274BS - 1	LCSW	15292	3.4454	-7470A	PPB	07 Mar 2023 14:50:49	U	SMPL	
O1798-03 - 1	TOTE-COMP-1	195	0.0214	-7470A	PPB	07 Mar 2023 14:53:05	U	SMPL	
O1798-03DUP - 1	TOTE-COMP-1DUP	158	0.0130	-7470A	PPB	07 Mar 2023 14:58:06	U	SMPL	
O1798-03MS - 1	TOTE-COMP-1MS	15097	3.4012	-7470A	PPB	07 Mar 2023 15:00:22	U	SMPL	
O1798-03MSD - 1	TOTE-COMP-1MSD	15486	3.4894	-7470A	PPB	07 Mar 2023 15:02:38	U	SMPL	
CCV31 - 1	CCV31	22829	5.1548	-7470A	PPB	07 Mar 2023 15:04:55	U	SMPL	
CCB31 - 1	CCB31	-40	-0.0319	-7470A	PPB	07 Mar 2023 15:07:11	U	SMPL	
O1803-08 - 1	1711	2784	0.6086	-7470A	PPB	07 Mar 2023 15:09:27	U	SMPL	
O1812-01 - 1	20230161-SITE-WATER	20	-0.0183	-7470A	PPB	07 Mar 2023 15:11:42	U	SMPL	
O1816-01 - 1	2321	854	0.1709	-7470A	PPB	07 Mar 2023 15:13:58	U	SMPL	
O1816-07 - 1	662	139	0.0087	-7470A	PPB	07 Mar 2023 15:16:14	U	SMPL	
PB151252TB - 1		-19	-0.0271	-7470A	PPB	07 Mar 2023 15:18:32	U	SMPL	
O1825-08LX5 - 1		-29	-0.0294	-7470A	PPB	07 Mar 2023 15:20:48	U	SMPL	
O1825-08A - 1		18908	4.2655	-7470A	PPB	07 Mar 2023 15:23:05	U	SMPL	
O1798-03LX5 - 1		-24	-0.0282	-7470A	PPB	07 Mar 2023 15:25:22	U	SMPL	
O1798-03A - 1		19063	4.3007	-7470A	PPB	07 Mar 2023 15:27:40	U	SMPL	
CCV32 - 1	CCV32	22131	4.9965	-7470A	PPB	07 Mar 2023 15:29:56	U	SMPL	
CCB32 - 1	CCB32	-13	-0.0257	-7470A	PPB	07 Mar 2023 15:32:11	U	SMPL	