

LB124401


SFAM01.1

INSTRUMENT ID: EV1

Linear

 μ Abs.:

52951



A= 0.0000e+000

B= 1.8926e-004 slope

C= -1.9941e-002 y-intercept

Rho= 0.9999811

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.012	-0.012	42	0.000	42					
0.05	0.050					185	0				
0.20	0.200	0.224	0.024	1291	0.0 %	1291					
2.50	2.500	2.467	-0.033	13142	0.0 %	13142					
5.00	5.000	5.033	0.033	26698	0.0 %	26698					
7.50	7.500	7.486	-0.014	39659	0.0 %	39659					
10.0	10.000	10.002	0.002	52951	0.0 %	52951					

12
11
10
00

LB124401

INSTRUMENT ID :CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	42	-		0 SFAM01.1	PPB	3/8/2023 7:59	Std
	0.2 S01	1291	-		0.2 SFAM01.1	PPB	3/8/2023 8:01	Std
	2.5 S02	13142	-		2.5 SFAM01.1	PPB	3/8/2023 8:04	Std
	5 S03	26698	-		5 SFAM01.1	PPB	3/8/2023 8:06	Std
	7.5 S04	39659	-		7.5 SFAM01.1	PPB	3/8/2023 8:08	Std
	10 S05	52951	-		10 SFAM01.1	PPB	3/8/2023 8:10	Std
ICV074	ICV074	22098	4.1623 -		SFAM01.1	PPB	3/8/2023 8:13	SMPL
ICB074	ICB074	-21	-0.0239 -		SFAM01.1	PPB	3/8/2023 8:15	SMPL
CCV033	CCV033	25961	4.8934 -		SFAM01.1	PPB	3/8/2023 8:18	SMPL
CCB033	CCB033	-17	-0.0232 -		SFAM01.1	PPB	3/8/2023 8:20	SMPL
PB151291BL	PBW291	55	-0.0095 -		SFAM01.1	PPB	3/8/2023 8:24	SMPL
O1731-01	YBKZ2	4151	0.7657 -		SFAM01.1	PPB	3/8/2023 8:27	SMPL
O1731-02	YBKZ2D	4881	0.9038 -		SFAM01.1	PPB	3/8/2023 8:29	SMPL
O1731-03	YBKZ2S	10116	1.8946 -		SFAM01.1	PPB	3/8/2023 8:31	SMPL
O1731-04	YBKZ4	14937	2.807 -		SFAM01.1	PPB	3/8/2023 8:33	SMPL
O1731-05	YBKZ6	15541	2.9213 -		SFAM01.1	PPB	3/8/2023 8:36	SMPL
O1731-06	YBKZ8	16689	3.1386 -		SFAM01.1	PPB	3/8/2023 8:38	SMPL
O1731-07	YBL18	6900	1.2859 -		SFAM01.1	PPB	3/8/2023 8:40	SMPL
O1731-08	YBL20	11600	2.1755 -		SFAM01.1	PPB	3/8/2023 8:43	SMPL
O1731-09	YBL22	31220	5.8887 -		SFAM01.1	PPB	3/8/2023 8:45	SMPL
O1731-10	YBL24	31528	5.947 -		SFAM01.1	PPB	3/8/2023 8:47	SMPL
O1731-11	YBL26	26304	4.9583 -		SFAM01.1	PPB	3/8/2023 8:49	SMPL
O1731-12	YBL28	23844	4.4928 -		SFAM01.1	PPB	3/8/2023 8:52	SMPL
O1731-13	YBL30	18805	3.5391 -		SFAM01.1	PPB	3/8/2023 8:54	SMPL
PB151292BL	PBW292	-33	-0.0262 -		SFAM01.1	PPB	3/8/2023 8:56	SMPL
O1732-01	YBKW0	26630	5.02 -		SFAM01.1	PPB	3/8/2023 8:58	SMPL
O1732-02	YBKW2	79076	14.9459 -		SFAM01.1	PPB	3/8/2023 9:01	SMPL
O1732-03	YBKW4	96551	18.2532 -		SFAM01.1	PPB	3/8/2023 9:03	SMPL
O1732-04	YBKW6	93343	17.6461 -		SFAM01.1	PPB	3/8/2023 9:05	SMPL
O1732-05	YBKX2	4182	0.7715 -		SFAM01.1	PPB	3/8/2023 9:07	SMPL
CCV034	CCV034	26653	5.0244 -		SFAM01.1	PPB	3/8/2023 9:10	SMPL
CCB034	CCB034	-23	-0.0243 -		SFAM01.1	PPB	3/8/2023 9:12	SMPL
O1732-06	YBKX4	66622	12.5889 -		SFAM01.1	PPB	3/8/2023 9:14	SMPL
O1732-07	YBKX6	76147	14.3916 -		SFAM01.1	PPB	3/8/2023 9:17	SMPL
O1732-08	YBKX8	13966	2.6233 -		SFAM01.1	PPB	3/8/2023 9:19	SMPL
O1732-09	YBKY0	17672	3.3246 -		SFAM01.1	PPB	3/8/2023 9:21	SMPL
O1732-10	YBKY4	4184	0.7719 -		SFAM01.1	PPB	3/8/2023 9:23	SMPL
O1732-11	YBKY6	5034	0.9328 -		SFAM01.1	PPB	3/8/2023 9:26	SMPL
O1732-12	YBKY8	2829	0.5155 -		SFAM01.1	PPB	3/8/2023 9:28	SMPL
O1732-13	YBKZ0	2117	0.3807 -		SFAM01.1	PPB	3/8/2023 9:30	SMPL
O1732-14	YBL02	9513	1.7805 -		SFAM01.1	PPB	3/8/2023 9:33	SMPL
O1732-15	YBL04	11919	2.2358 -		SFAM01.1	PPB	3/8/2023 9:35	SMPL
O1732-16	YBL06	13319	2.5008 -		SFAM01.1	PPB	3/8/2023 9:37	SMPL
O1732-17	YBL06D	12900	2.4215 -		SFAM01.1	PPB	3/8/2023 9:39	SMPL
O1732-18	YBL06S	17726	3.3349 -		SFAM01.1	PPB	3/8/2023 9:42	SMPL
O1732-19	YBL10	8930	1.6701 -		SFAM01.1	PPB	3/8/2023 9:44	SMPL

LB124401

INSTRUMENT ID :CV1

O1732-20	YBL12	9474	1.7731 -	SFAM01.1	PPB	3/8/2023 9:46 SMPL
O1732-21	YBL14	706	0.1137 -	SFAM01.1	PPB	3/8/2023 9:48 SMPL
O1732-22	YBL52	1230	0.2128 -	SFAM01.1	PPB	3/8/2023 9:51 SMPL
PB151293BL	PBW293	33	-0.0137 -	SFAM01.1	PPB	3/8/2023 9:57 SMPL
O1733-01	YBKZ3	616	0.0966 -	SFAM01.1	PPB	3/8/2023 9:59 SMPL
O1733-02	YBKZ3D	640	0.1012 -	SFAM01.1	PPB	3/8/2023 10:02 SMPL
CCV035	CCV035	26215	4.9415 -	SFAM01.1	PPB	3/8/2023 10:04 SMPL
CCB035	CCB035	-5	-0.0209 -	SFAM01.1	PPB	3/8/2023 10:06 SMPL
O1733-03	YBKZ3S	6504	1.211 -	SFAM01.1	PPB	3/8/2023 10:08 SMPL
O1733-04	YBKZ5	673	0.1074 -	SFAM01.1	PPB	3/8/2023 10:11 SMPL
O1733-05	YBKZ7	747	0.1214 -	SFAM01.1	PPB	3/8/2023 10:13 SMPL
O1733-06	YBKZ9	764	0.1247 -	SFAM01.1	PPB	3/8/2023 10:15 SMPL
O1733-07	YBL19	783	0.1282 -	SFAM01.1	PPB	3/8/2023 10:17 SMPL
O1733-08	YBL21	813	0.1339 -	SFAM01.1	PPB	3/8/2023 10:20 SMPL
O1733-09	YBL23	787	0.129 -	SFAM01.1	PPB	3/8/2023 10:22 SMPL
O1733-10	YBL25	607	0.0949 -	SFAM01.1	PPB	3/8/2023 10:24 SMPL
O1733-11	YBL27	990	0.1674 -	SFAM01.1	PPB	3/8/2023 10:27 SMPL
O1733-12	YBL29	732	0.1186 -	SFAM01.1	PPB	3/8/2023 10:29 SMPL
O1733-13	YBL31	699	0.1124 -	SFAM01.1	PPB	3/8/2023 10:31 SMPL
O1733-14	YBKY7	739	0.1199 -	SFAM01.1	PPB	3/8/2023 10:33 SMPL
O1733-15	YBKY9	736	0.1194 -	SFAM01.1	PPB	3/8/2023 10:36 SMPL
CCV036	CCV036	26329	4.9631 -	SFAM01.1	PPB	3/8/2023 10:38 SMPL
CCB036	CCB036	-42	-0.0279 -	SFAM01.1	PPB	3/8/2023 10:40 SMPL
O1732-03DLX5		19064	3.5881 -	SFAM01.1	PPB	3/8/2023 10:45 SMPL
O1732-06DLX2		30072	5.6715 -	SFAM01.1	PPB	3/8/2023 10:55 SMPL
O1732-07DLX2		36069	6.8064 -	SFAM01.1	PPB	3/8/2023 11:01 SMPL
O1732-02DLX2		35586	6.715 -	SFAM01.1	PPB	3/8/2023 11:06 SMPL
O1732-04DLX5		19467	3.6644 -	SFAM01.1	PPB	3/8/2023 11:09 SMPL
CCV037	CCV037	25781	4.8593 -	SFAM01.1	PPB	3/8/2023 11:12 SMPL
CCB037	CCB037	-25	-0.0247 -	SFAM01.1	PPB	3/8/2023 11:14 SMPL