



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB124406

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/08/2023	12:35
2	CAL2	1	Water	NA	NA	20.7	7.01	03/08/2023	12:36
3	CAL3	1	Water	NA	NA	20.7	10.03	03/08/2023	12:38
4	ICV	1	Water	NA	NA	20.8	7.02	03/08/2023	12:40
5	CCV1	1	Water	NA	NA	20.7	2.01	03/08/2023	12:45
6	O1798-03	1	Water	NA	NA	23.2	5.34	03/08/2023	12:47
7	O1798-03DUP	1	Water	NA	NA	23.3	5.35	03/08/2023	12:48
8	O1840-01	1	Water	NA	NA	20.4	4.33	03/08/2023	12:50
9	O1840-02	1	Water	NA	NA	20.2	4.45	03/08/2023	12:55
10	O1840-03	1	Water	NA	NA	20.1	4.33	03/08/2023	13:00
11	CCV2	1	Water	NA	NA	20.7	12.02	03/08/2023	13:02

1312406

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O1840 WorkList ID : 168004 Department : Wet-Chemistry Date : 03-08-2023 12:21:43

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
03/03/2023	Water	O1798-03	pH A	Cool 4 deg C	PSEG03	E31	TOTE-COMP-1	03/03/2023 9040C
02/28/2023	Water	O1840-01	pH A	Cool 4 deg C	STER09	N11	CELL-A-POOL-WATER	02/28/2023 9040C
02/28/2023	Water	O1840-02	pH A	Cool 4 deg C	STER09	N11	CELL-B/C-POOL-WATER	02/28/2023 9040C
02/28/2023	Water	O1840-03	pH A	Cool 4 deg C	STER09	N11	CELL-D-POOL-WATER	02/28/2023 9040C

Date/Time 03/08/23 12:30
Raw Sample Received by: 10/001
Raw Sample Relinquished by: 12/01

Date/Time 03/08/23 13:35
Raw Sample Received by: 12/01
Raw Sample Relinquished by: 10/001